

Data Path : W:\HPCHEM1\MSVOA U\DATA\VU050818\
 Data File : VU023737.D
 Acq On : 08 May 2018 15:12
 Operator : MD/SY
 Sample : J2721-13
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Quant Time: May 09 03:59:31 2018
 Quant Method : W:\HPCHEM1\MSVOA_U\METHOD\SOMULM050818WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed May 09 03:02:40 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	356617	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	339923	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.49	152	166222	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	127834	45.83	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	91.66%
7) Chloroethane-d5	1.68	69	98123	47.10	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	94.20%
11) 1,1-Dichloroethene-d2	2.27	63	183686	35.54	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.08%
21) 2-Butanone-d5	4.18	46	209067	101.84	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	101.84%
24) Chloroform-d	4.65	84	211181	45.53	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.06%
26) 1,2-Dichloroethane-d4	5.31	65	143283	47.75	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.50%
32) Benzene-d6	5.35	84	461684	47.41	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.82%
36) 1,2-Dichloropropane-d6	6.34	67	154060	48.37	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.74%
41) Toluene-d8	7.57	98	417192	46.17	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.34%
43) trans-1,3-Dichloropropene-	7.85	79	63840	46.33	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.66%
47) 2-Hexanone-d5	8.31	63	149864	105.32	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	105.32%
57) 1,1,2,2-Tetrachloroethane-	10.44	84	202528	46.93	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	93.86%
64) 1,2-Dichlorobenzene-d4	11.87	152	167524	48.66	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.32%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA U\DATA\VU050818\
Data File : VU023737.D
Acq On : 08 May 2018 15:12
Operator : MD/SY
Sample : J2721-13
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 13 Sample Multiplier: 1

Quant Time: May 09 03:59:31 2018
Quant Method : W:\HPCHEM1\MSVOA_U\METHOD\SOMULM050818WMA.M
Quant Title : VOC Analysis
QLast Update : Wed May 09 03:02:40 2018
Response via : Initial Calibration

