

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050823\
 Data File : VU054075.D
 Acq On : 08 May 2023 13:00
 Operator : JC/MD
 Sample : VU0508MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK091

Quant Time: May 09 04:38:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM042923WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat May 06 01:14:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.244	114	216687	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.411	117	201293	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.806	152	95822	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	74444	45.425	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	90.860%
7) Chloroethane-d5	1.906	69	64800	60.237	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	120.480%
11) 1,1-Dichloroethene-d2	2.562	63	110398	37.981	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	75.960%
21) 2-Butanone-d5	4.617	46	112718	120.595	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	120.590%
24) Chloroform-d	5.057	84	160953	53.475	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	106.940%
26) 1,2-Dichloroethane-d4	5.694	65	107979	55.874	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	111.740%
32) Benzene-d6	5.719	84	318875	53.818	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	107.640%
36) 1,2-Dichloropropane-d6	6.684	67	103614	55.394	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	110.780%
41) Toluene-d8	7.893	98	282229	52.760	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	105.520%
43) trans-1,3-Dichloroprop...	8.176	79	48667	52.575	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	105.160%
47) 2-Hexanone-d5	8.626	63	72335	107.969	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	107.970%
56) 1,1,2,2-Tetrachloroeth...	10.751	84	141563	57.457	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	114.920%
66) 1,2-Dichlorobenzene-d4	12.189	152	110332	58.941	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	117.880%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050823\
Data File : VU054075.D
Acq On : 08 May 2023 13:00
Operator : JC/MD
Sample : VU0508MBL01
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK091

Quant Time: May 09 04:38:43 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM042923WMA.M
Quant Title : VOC Analysis
QLast Update : Sat May 06 01:14:00 2023
Response via : Initial Calibration

