

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050823\
 Data File : VU054106.D
 Acq On : 09 May 2023 01:49
 Operator : JC/MD
 Sample : 02663-16
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 JREK6

Quant Time: May 09 06:13:53 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM042923WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat May 06 01:14:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.243	114	215891	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.410	117	203629	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.806	152	101648	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	57889	35.454	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	70.900%	
7) Chloroethane-d5	1.906	69	51468	48.020	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	96.040%	
11) 1,1-Dichloroethene-d2	2.562	63	91197	31.491	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	62.980%	
21) 2-Butanone-d5	4.616	46	96004	103.092	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	103.090%	
24) Chloroform-d	5.057	84	141337	47.131	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	94.260%	
26) 1,2-Dichloroethane-d4	5.694	65	93846	48.740	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	97.480%	
32) Benzene-d6	5.722	84	269275	44.925	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	89.860%	
36) 1,2-Dichloropropane-d6	6.684	67	89625	47.366	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	94.740%	
41) Toluene-d8	7.893	98	238318	44.040	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	88.080%	
43) trans-1,3-Dichloroprop...	8.173	79	42147	45.009	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	90.020%	
47) 2-Hexanone-d5	8.626	63	61166	90.251	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	90.250%	
56) 1,1,2,2-Tetrachloroeth...	10.751	84	124768	50.059	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	100.120%	
66) 1,2-Dichlorobenzene-d4	12.188	152	101342	51.035	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	102.080%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU050823\
Data File : VU054106.D
Acq On : 09 May 2023 01:49
Operator : JC/MD
Sample : 02663-16
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
JREK6

Quant Time: May 09 06:13:53 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM042923WMA.M
Quant Title : VOC Analysis
QLast Update : Sat May 06 01:14:00 2023
Response via : Initial Calibration

