

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU051021\
 Data File : VU043634.D
 Acq On : 10 May 2021 13:30
 Operator : SY/MD
 Sample : M2271-11
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EW6D6

Quant Time: May 11 03:59:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR050321WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue May 11 03:58:06 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.256	114	97300	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.423	117	95271	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.819	152	43723	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.594	65	29588	3.78	ug/L	0.00
Spiked Amount 5.000	Range 40 - 130		Recovery = 75.60%			
7) Chloroethane-d5	1.912	69	23060	4.02	ug/L	0.00
Spiked Amount 5.000	Range 65 - 130		Recovery = 80.40%			
11) 1,1-Dichloroethene-d2	2.565	65	14116	3.68	ug/L	0.00
Spiked Amount 5.000	Range 60 - 125		Recovery = 73.60%			
20) 2-Butanone-d5	4.684	46	127490	52.49	ug/L	0.04
Spiked Amount 50.000	Range 40 - 130		Recovery = 104.98%			
24) Chloroform-d	5.073	84	67347	4.50	ug/L	0.00
Spiked Amount 5.000	Range 70 - 125		Recovery = 90.00%			
26) 1,2-Dichloroethane-d4	5.710	65	42193	4.68	ug/L	0.00
Spiked Amount 5.000	Range 70 - 130		Recovery = 93.60%			
32) Benzene-d6	5.735	84	120048	4.66	ug/L	0.00
Spiked Amount 5.000	Range 70 - 125		Recovery = 93.20%			
36) 1,2-Dichloropropane-d6	6.700	67	44542	4.93	ug/L	0.00
Spiked Amount 5.000	Range 60 - 140		Recovery = 98.60%			
41) Toluene-d8	7.906	98	103049	4.35	ug/L	0.00
Spiked Amount 5.000	Range 70 - 130		Recovery = 87.00%			
43) trans-1,3-Dichloroprop...	8.186	79	16631	4.56	ug/L	0.00
Spiked Amount 5.000	Range 55 - 130		Recovery = 91.20%			
46) 2-Hexanone-d5	8.649	63	88587	57.93	ug/L	0.00
Spiked Amount 50.000	Range 45 - 130		Recovery = 115.86%			
56) 1,1,2,2-Tetrachloroeth...	10.764	84	36380	5.02	ug/L	0.00
Spiked Amount 5.000	Range 65 - 120		Recovery = 100.40%			
66) 1,2-Dichlorobenzene-d4	12.201	152	38074	4.88	ug/L	0.00
Spiked Amount 5.000	Range 80 - 120		Recovery = 97.60%			
Target Compounds						Qvalue
16) Methylene chloride	3.047	84	7838	0.90	ug/L	97
42) Toluene	7.977	91	2350	0.09	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU051021\
Data File : VU043634.D
Acq On : 10 May 2021 13:30
Operator : SY/MD
Sample : M2271-11
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EW6D6

Quant Time: May 11 03:59:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR050321WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue May 11 03:58:06 2021
Response via : Initial Calibration

