

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU051021\
 Data File : VU043650.D
 Acq On : 10 May 2021 19:55
 Operator : SY/MD
 Sample : M2301-01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BG6J1

Quant Time: May 11 04:01:51 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR050321WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue May 11 03:58:06 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.259	114	93656	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.426	117	94533	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.819	152	41926	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	24977	3.32	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	66.40%
7) Chloroethane-d5	1.919	69	21753	3.94	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	78.80%
11) 1,1-Dichloroethene-d2	2.571	65	12947	3.50	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	70.00%
20) 2-Butanone-d5	4.642	46	104070	44.51	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	89.02%
24) Chloroform-d	5.073	84	61271	4.25	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.00%
26) 1,2-Dichloroethane-d4	5.713	65	38888	4.48	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.60%
32) Benzene-d6	5.735	84	108884	4.26	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.20%
36) 1,2-Dichloropropane-d6	6.700	67	40559	4.52	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	90.40%
41) Toluene-d8	7.906	98	88124	3.75	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	75.00%
43) trans-1,3-Dichloroprop...	8.185	79	15679	4.33	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	86.60%
46) 2-Hexanone-d5	8.642	63	84268	55.54	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	111.08%
56) 1,1,2,2-Tetrachloroeth...	10.764	84	34565	4.81	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.20%
66) 1,2-Dichlorobenzene-d4	12.201	152	34780	4.65	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	93.00%
Target Compounds						
					Qvalue	
19) 1,1-Dichloroethane	3.880	63	2852	0.22	ug/L	88
22) cis-1,2-Dichloroethene	4.681	96	1016	0.16	ug/L #	90
25) Chloroform	5.099	83	9386	0.68	ug/L	97
34) Trichloroethene	6.552	95	1135	0.18	ug/L #	81
47) Tetrachloroethene	8.562	164	5168	1.10	ug/L	94


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU051021\
Data File : VU043650.D
Acq On : 10 May 2021 19:55
Operator : SY/MD
Sample : M2301-01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BG6J1

Quant Time: May 11 04:01:51 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR050321WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue May 11 03:58:06 2021
Response via : Initial Calibration









#19

1,1-Dichloroethane
Concen: 0.22 ug/L
RT: 3.880 min Scan# 790
Delta R.T. -0.000 min
Lab File: VU043650.D
Acq: 10 May 2021 19:55

Instrument :
MSVOA_U
ClientSampleId :
BG6J1




Tgt Ion: 63	Resp: 2852
Ion Ratio	Lower Upper
63 100	
65 25.0	22.7 42.1
83 9.8	9.4 17.4




#22

cis-1,2-Dichloroethene
Concen: 0.16 ug/L
RT: 4.681 min Scan# 1039
Delta R.T. 0.003 min
Lab File: VU043650.D
Acq: 10 May 2021 19:55

Tgt Ion: 96	Resp: 1016		
Ion Ratio	Lower	Upper	
96	100		
61	150.9	102.8	190.8
98	41.6	42.8	79.6#









#25

Chloroform
Concen: 0.68 ug/L
RT: 5.099 min Scan# 1169
Delta R.T. -0.000 min
Lab File: VU043650.D
Acq: 10 May 2021 19:55

Instrument :
MSVOA_U
ClientSampleId :
BG6J1




Tgt Ion: 83 Resp: 9386
Ion Ratio Lower Upper
83 100
85 61.8 45.1 83.9




#34

Trichloroethene
Concen: 0.18 ug/L
RT: 6.552 min Scan# 1621
Delta R.T. 0.003 min
Lab File: VU043650.D
Acq: 10 May 2021 19:55

Tgt Ion: 95 Resp: 1135
Ion Ratio Lower Upper
95 100
97 51.2 46.5 86.5
132 64.0 64.1 119.1#
130 86.3 67.3 125.1





#47
Tetrachloroethene
Concen: 1.10 ug/L
RT: 8.562 min Scan# 2246
Delta R.T. -0.000 min
Lab File: VU043650.D
Acq: 10 May 2021 19:55

Instrument :
MSVOA_U
ClientSampleId :
BG6J1

Tgt Ion:164 Resp: 5168

Ion	Ratio	Lower	Upper
164	100		
129	103.0	62.5	116.1
131	94.3	64.9	120.5
166	130.9	89.0	165.2

