

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU051022\
 Data File : VU048484.D
 Acq On : 10 May 2022 12:30
 Operator : SY/MD
 Sample : N2736-05
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EXEW0

Quant Time: May 11 05:21:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM050622WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed May 11 05:17:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	275237	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	273388	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	137392	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	86336	36.423	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	72.840%	
7) Chloroethane-d5	1.912	69	68961	44.577	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	89.160%	
11) 1,1-Dichloroethene-d2	2.568	63	136767	32.420	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	64.840%	
21) 2-Butanone-d5	4.620	46	165613	82.039	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	82.040%	
24) Chloroform-d	5.063	84	189974	46.142	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	92.280%	
26) 1,2-Dichloroethane-d4	5.703	65	137202	51.265	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	102.520%	
32) Benzene-d6	5.729	84	364177	43.836	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	87.680%	
36) 1,2-Dichloropropane-d6	6.690	67	115012	42.454	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	84.900%	
41) Toluene-d8	7.896	98	320920	43.678	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	87.360%	
43) trans-1,3-Dichloroprop...	8.179	79	43571	34.943	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	69.880%	
47) 2-Hexanone-d5	8.632	63	120035	91.543	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	91.540%	
56) 1,1,2,2-Tetrachloroeth...	10.758	84	169108	40.980	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	81.960%	
66) 1,2-Dichlorobenzene-d4	12.192	152	131190	48.221	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	96.440%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU051022\
Data File : VU048484.D
Acq On : 10 May 2022 12:30
Operator : SY/MD
Sample : N2736-05
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EXEW0

Quant Time: May 11 05:21:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM050622WMA.M
Quant Title : VOC Analysis
QLast Update : Wed May 11 05:17:27 2022
Response via : Initial Calibration

