

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU051023\
 Data File : VU054147.D
 Acq On : 10 May 2023 16:00
 Operator : JC/MD
 Sample : 02694-23
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: May 11 05:13:15 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM042923WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat May 06 01:14:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.243	114	257597	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.410	117	245992	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.806	152	126168	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	76493	39.263	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	78.520%
7) Chloroethane-d5	1.906	69	63135	49.368	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	98.740%
11) 1,1-Dichloroethene-d2	2.562	63	110064	31.852	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	63.700%
21) 2-Butanone-d5	4.616	46	99196	89.273	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	89.270%
24) Chloroform-d	5.057	84	161481	45.130	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.260%
26) 1,2-Dichloroethane-d4	5.694	65	105845	46.072	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.140%
32) Benzene-d6	5.719	84	308700	42.633	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.260%
36) 1,2-Dichloropropane-d6	6.684	67	100227	43.847	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	87.700%
41) Toluene-d8	7.893	98	279942	42.823	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	85.640%
43) trans-1,3-Dichloroprop...	8.172	79	50731	44.846	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	89.700%
47) 2-Hexanone-d5	8.626	63	66507	81.232	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	81.230%
56) 1,1,2,2-Tetrachloroeth...	10.751	84	137709	45.736	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.480%
66) 1,2-Dichlorobenzene-d4	12.188	152	121429	49.267	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.540%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU051023\
Data File : VU054147.D
Acq On : 10 May 2023 16:00
Operator : JC/MD
Sample : 02694-23
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: May 11 05:13:15 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM042923WMA.M
Quant Title : VOC Analysis
QLast Update : Sat May 06 01:14:00 2023
Response via : Initial Calibration

