


Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU051023\
Data File : VU054135.D
Acq On : 10 May 2023 09:58
Operator : JC/MD
Sample : BFB189
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFB189

Integration File: LSCINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM042923WMA.M
Title : VOC Analysis
Last Update : Sat May 06 01:14:00 2023



AutoFind: Scans 2888, 2889, 2890; Background Corrected with Scan 2877

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	16.4	6299	PASS
75	95	30	80	50.0	19235	PASS
95	95	100	100	100.0	38480	PASS
96	95	5	9	7.2	2784	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	84.7	32595	PASS
175	174	5	9	8.0	2597	PASS
176	174	95	101	97.7	31861	PASS
177	176	5	9	6.9	2194	PASS

m/z	Abundance
39.80	135.0
43.90	221.0

Instrument :
MSVOA_U
ClientSampleId :
BFB189

m/z	Abundance
35.90	219.0
36.90	1437.0
38.00	1358.0
39.00	519.0
43.90	377.0
44.90	292.0
47.00	358.0
47.90	271.0
48.90	1442.0
50.00	6415.0
51.00	1923.0
52.00	122.0
54.80	120.0
55.90	558.0
57.00	875.0
59.90	317.0
61.00	1784.0
62.00	2046.0
63.00	1393.0
63.90	208.0
65.00	202.0
68.00	3622.0
69.00	3854.0
69.90	317.0
71.90	215.0
73.00	1701.0
74.00	6534.0
75.00	19304.0
76.00	1583.0
76.80	170.0
78.10	106.0
78.90	1177.0
79.90	417.0
80.80	1385.0
82.00	274.0
86.90	1320.0
87.90	1488.0
91.00	155.0
92.00	1333.0
93.00	1883.0
94.00	4426.0
95.00	38272.0
96.00	2892.0
104.00	218.0
105.90	214.0
115.80	156.0
116.90	277.0
117.70	204.0
118.90	190.0
127.80	172.0
140.70	371.0
142.90	582.0
171.70	145.0
174.00	31480.0
175.00	2600.0
175.90	31056.0
176.90	2079.0

Instrument :
MSVOA_U
ClientSampleId :
BFB189

m/z	Abundance
35.90	237.0
36.90	1440.0
38.00	1427.0
38.90	564.0
43.90	379.0
44.90	357.0
47.00	338.0
47.80	271.0
48.90	1447.0
50.00	6422.0
51.00	1987.0
52.00	150.0
54.90	134.0
55.90	565.0
57.00	944.0
59.90	411.0
61.00	1866.0
62.00	2003.0
63.00	1381.0
63.90	237.0
65.00	232.0
68.00	3828.0
69.00	3992.0
69.90	313.0
71.80	262.0
72.90	1636.0
74.00	6450.0
75.00	19792.0
76.00	1642.0
76.80	218.0
78.90	1205.0
79.90	472.0
80.80	1347.0
82.00	266.0
87.00	1590.0
87.90	1571.0
90.90	186.0
92.00	1433.0
93.00	2084.0
94.00	4403.0
95.00	39872.0
96.00	2881.0
104.00	195.0
105.90	169.0
115.80	190.0
116.90	258.0
117.70	199.0
118.90	210.0
127.80	198.0
129.80	106.0
140.70	361.0
142.90	603.0
154.70	123.0
157.00	131.0
171.80	209.0
173.90	33664.0
175.00	2721.0
175.90	32952.0
177.00	2317.0

Instrument :
MSVOA_U
ClientSampleId :
BFB189

m/z	Abundance
35.90	246.0
37.00	1367.0
38.00	1276.0
38.90	566.0
43.90	339.0
44.90	347.0
47.00	300.0
47.90	251.0
49.00	1311.0
50.00	6061.0
51.00	1870.0
52.00	127.0
55.90	536.0
57.00	943.0
59.90	420.0
61.00	1827.0
62.00	1766.0
62.90	1329.0
63.90	227.0
65.00	217.0
68.00	3721.0
69.00	3707.0
69.90	294.0
71.80	295.0
72.90	1506.0
74.00	5821.0
75.00	18608.0
76.00	1662.0
76.80	239.0
78.90	1094.0
79.90	441.0
80.90	1189.0
81.90	227.0
87.00	1740.0
87.90	1463.0
90.90	163.0
91.90	1372.0
92.90	1964.0
94.00	4102.0
95.00	37296.0
96.00	2580.0
104.00	180.0
105.80	116.0
115.80	188.0
116.90	253.0
117.60	141.0
118.80	206.0
127.80	173.0
129.80	118.0
136.90	103.0
140.80	338.0
142.90	551.0
154.70	128.0
171.80	265.0
173.90	32640.0
175.00	2469.0
175.90	31576.0
177.00	2185.0

Instrument :
MSVOA_U
ClientSampleId :
BFB189