

Data Path : W:\HPCHEM1\MSVOA U\DATA\VU051118\
 Data File : VU023794.D
 Acq On : 11 May 2018 12:21
 Operator : MD/SY
 Sample : J2777-11
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Quant Time: May 12 02:17:49 2018
 Quant Method : W:\HPCHEM1\MSVOA U\METHOD\SOMUTR051018WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat May 12 02:07:11 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	138513	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	141403	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.49	152	61293	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	56496	5.12	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	102.40%
7) Chloroethane-d5	1.68	69	44388	5.45	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	109.00%
11) 1,1-Dichloroethene-d2	2.27	63	71914	3.71	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	74.20%
20) 2-Butanone-d5	4.22	46	131477	51.09	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	102.18%
24) Chloroform-d	4.65	84	83147	4.64	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.80%
26) 1,2-Dichloroethane-d4	5.32	65	47542	4.93	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.60%
32) Benzene-d6	5.35	84	182985	4.72	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.40%
36) 1,2-Dichloropropane-d6	6.34	67	60965	4.84	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.80%
41) Toluene-d8	7.57	98	152507	4.26	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.20%
43) trans-1,3-Dichloropropene-	7.86	79	18625	4.42	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	88.40%
46) 2-Hexanone-d5	8.32	63	94480	49.29	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	98.58%
57) 1,1,2,2-Tetrachloroethane-	10.44	84	43871	4.80	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.00%
64) 1,2-Dichlorobenzene-d4	11.87	152	59945	5.11	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA U\DATA\VU051118\
Data File : VU023794.D
Acq On : 11 May 2018 12:21
Operator : MD/SY
Sample : J2777-11
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 4 Sample Multiplier: 1

Quant Time: May 12 02:17:49 2018
Quant Method : W:\HPCHEM1\MSVOA U\METHOD\SOMUTR051018WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat May 12 02:07:11 2018
Response via : Initial Calibration

