

Data Path : W:\HPCHEM1\MSVOA U\DATA\VU051118\
 Data File : VU023796.D
 Acq On : 11 May 2018 13:12
 Operator : MD/SY
 Sample : J2746-03
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Quant Time: May 12 02:27:18 2018
 Quant Method : W:\HPCHEM1\MSVOA U\METHOD\SOMUTR051018WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat May 12 02:07:11 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	129123	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	130154	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.49	152	56390	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	56102	5.45	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	109.00%
7) Chloroethane-d5	1.68	69	44868	5.90	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	118.00%
11) 1,1-Dichloroethene-d2	2.27	63	78047	4.32	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	86.40%
20) 2-Butanone-d5	4.22	46	126604	52.77	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	105.54%
24) Chloroform-d	4.65	84	80078	4.80	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.00%
26) 1,2-Dichloroethane-d4	5.32	65	46642	5.19	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.80%
32) Benzene-d6	5.35	84	173370	4.86	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.20%
36) 1,2-Dichloropropane-d6	6.34	67	58128	5.02	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.40%
41) Toluene-d8	7.57	98	142389	4.32	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.40%
43) trans-1,3-Dichloropropene-	7.86	79	17574	4.53	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.60%
46) 2-Hexanone-d5	8.32	63	89794	50.90	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	101.80%
57) 1,1,2,2-Tetrachloroethane-	10.44	84	43301	5.14	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	102.80%
64) 1,2-Dichlorobenzene-d4	11.87	152	57710	5.34	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	106.80%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA U\DATA\VU051118\
Data File : VU023796.D
Acq On : 11 May 2018 13:12
Operator : MD/SY
Sample : J2746-03
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 6 Sample Multiplier: 1

Quant Time: May 12 02:27:18 2018
Quant Method : W:\HPCHEM1\MSVOA U\METHOD\SOMUTR051018WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat May 12 02:07:11 2018
Response via : Initial Calibration

