

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU051122\
 Data File : VU048511.D
 Acq On : 11 May 2022 12:43
 Operator : SY/MD
 Sample : N2736-15
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EXEX1

Quant Time: May 12 03:42:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM050622WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu May 12 03:40:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	267901	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	265302	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	139330	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	60116	26.056	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	52.120%#
7) Chloroethane-d5	1.912	69	53295	35.394	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	70.780%
11) 1,1-Dichloroethene-d2	2.568	63	110846	26.995	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	53.980%#
21) 2-Butanone-d5	4.620	46	160051	81.455	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	81.450%
24) Chloroform-d	5.063	84	177440	44.277	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.560%
26) 1,2-Dichloroethane-d4	5.703	65	130578	50.126	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.260%
32) Benzene-d6	5.729	84	323879	40.174	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	80.340%
36) 1,2-Dichloropropane-d6	6.690	67	101540	38.623	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	77.240%
41) Toluene-d8	7.899	98	283857	39.811	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	79.620%#
43) trans-1,3-Dichloroprop...	8.182	79	56592	46.769	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	93.540%
47) 2-Hexanone-d5	8.632	63	115027	90.398	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	90.400%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	161257	40.268	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	80.540%
66) 1,2-Dichlorobenzene-d4	12.192	152	129174	46.820	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.640%
Target Compounds						
					Qvalue	
13) Acetone	2.626	43	17068	11.522	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU051122\
Data File : VU048511.D
Acq On : 11 May 2022 12:43
Operator : SY/MD
Sample : N2736-15
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EXEX1

Quant Time: May 12 03:42:13 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM050622WMA.M
Quant Title : VOC Analysis
QLast Update : Thu May 12 03:40:41 2022
Response via : Initial Calibration

