

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU051122\
 Data File : VU048523.D
 Acq On : 11 May 2022 17:38
 Operator : SY/MD
 Sample : N2790-09
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BGP8

Quant Time: May 12 03:44:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM050622WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu May 12 03:40:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	265310	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	269665	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	142217	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	79321	34.715	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	69.440%
7) Chloroethane-d5	1.912	69	61591	41.303	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	82.600%
11) 1,1-Dichloroethene-d2	2.568	63	132163	32.500	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	65.000%
21) 2-Butanone-d5	4.620	46	154841	79.573	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	79.570%
24) Chloroform-d	5.063	84	196377	49.481	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.960%
26) 1,2-Dichloroethane-d4	5.703	65	139701	54.151	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	108.300%
32) Benzene-d6	5.729	84	359509	43.872	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.740%
36) 1,2-Dichloropropane-d6	6.694	67	110464	41.338	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	82.680%
41) Toluene-d8	7.899	98	330732	45.635	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.260%
43) trans-1,3-Dichloroprop...	8.179	79	59043	48.005	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	96.000%
47) 2-Hexanone-d5	8.632	63	112683	87.123	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	87.120%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	170291	41.836	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	83.680%
66) 1,2-Dichlorobenzene-d4	12.192	152	136978	48.641	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.280%
Target Compounds						
					Qvalue	
13) Acetone	2.626	43	9241	6.299	ug/L	94
25) Chloroform	5.086	83	15778	3.907	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU051122\
Data File : VU048523.D
Acq On : 11 May 2022 17:38
Operator : SY/MD
Sample : N2790-09
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BGP8

Quant Time: May 12 03:44:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM050622WMA.M
Quant Title : VOC Analysis
QLast Update : Thu May 12 03:40:41 2022
Response via : Initial Calibration

