

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU051219\
 Data File : VU031947.D
 Acq On : 12 May 2019 16:50
 Operator : JC/SP
 Sample : K2791-10
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F9D49

Quant Time: May 13 07:37:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR050819WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon May 13 07:00:03 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	464167	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	481039	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	174361	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	122846	3.88	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	77.60%
7) Chloroethane-d5	1.68	69	122223	4.68	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	93.60%
11) 1,1-Dichloroethene-d2	2.27	63	186956	2.74	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	54.80%#
20) 2-Butanone-d5	4.21	46	531913	43.07	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	86.14%
24) Chloroform-d	4.64	84	278055	4.29	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.80%
26) 1,2-Dichloroethane-d4	5.31	65	157214	3.99	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.80%
32) Benzene-d6	5.33	84	532129	4.32	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.40%
36) 1,2-Dichloropropane-d6	6.33	67	190858	4.22	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	84.40%
41) Toluene-d8	7.56	98	389304	3.70	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	74.00%
43) trans-1,3-Dichloropropene-	7.85	79	54247	3.42	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	68.40%
46) 2-Hexanone-d5	8.31	63	364515	45.42	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	90.84%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	168274	4.78	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.60%
64) 1,2-Dichlorobenzene-d4	11.86	152	149210	5.17	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.40%

Target Compounds

					Ovalue
13) Acetone	2.34	43	156169	18.513 ug/L	85
22) cis-1,2-Dichloroethene	4.22	96	44740	1.095 ug/L	83
34) Trichloroethene	6.19	95	15764	0.373 ug/L	92
47) Tetrachloroethene	8.22	164	122839	4.467 ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU051219\
Data File : VU031947.D
Acq On : 12 May 2019 16:50
Operator : JC/SP
Sample : K2791-10
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F9D49

Quant Time: May 13 07:37:20 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR050819WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon May 13 07:00:03 2019
Response via : Initial Calibration

