


Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU051220\
 Data File : VU038113.D
 Acq On : 12 May 2020 09:31
 Operator : JC/MD
 Sample : BFB21
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFB21

Integration File: rteint.p

Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR050720WMA.M
 Title : TRACE VOA SOM01.0
 Last Update : Wed May 13 05:52:50 2020



AutoFind: Scans 2895, 2896, 2897; Background Corrected with Scan 2882

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.1	30808	PASS
75	95	30	80	52.3	76250	PASS
95	95	100	100	100.0	145834	PASS
96	95	5	9	7.0	10185	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	73.4	107061	PASS
175	174	5	9	7.5	8003	PASS
176	174	95	101	95.5	102264	PASS
177	176	5	9	6.6	6733	PASS

m/z	Abundance
39.80	228.0
40.10	240.0
44.00	965.0

Instrument :
MSVOA_U
ClientSampleId :
BFB21

m/z	Abundance
35.90	1222.0
37.00	6525.0
38.00	6492.0
39.00	2801.0
40.00	469.0
43.00	110.0
44.00	1645.0
45.00	1322.0
45.80	173.0
47.00	1632.0
48.00	786.0
49.00	6241.0
50.00	29504.0
51.00	9041.0
52.00	566.0
55.00	250.0
56.00	1917.0
57.00	3810.0
57.80	115.0
60.00	1055.0
61.00	6954.0
62.00	7119.0
63.00	4960.0
64.00	491.0
65.00	357.0
67.10	391.0
68.00	15219.0
69.00	15862.0
70.00	1135.0
71.00	164.0
72.00	718.0
73.10	6385.0
74.00	24816.0
75.00	72680.0
76.00	6271.0
77.00	1071.0
78.00	622.0
78.90	3960.0
79.90	1185.0
80.90	3906.0
81.90	1055.0
83.00	135.0
86.10	272.0
87.00	6683.0
87.90	5920.0
90.80	398.0
92.00	3985.0
93.00	6106.0
94.00	15928.0
95.00	138624.0
96.00	9787.0
97.10	442.0
104.00	588.0
104.80	245.0
105.90	436.0
107.10	133.0
114.90	167.0
115.90	444.0
116.90	636.0
117.90	471.0
118.90	941.0
128.00	291.0
128.90	234.0
130.00	415.0
134.90	254.0
136.90	186.0
140.10	128.0
141.00	1013.0
141.80	139.0
142.90	1348.0
145.00	114.0
145.90	226.0
147.90	255.0
148.90	102.0
149.80	122.0
154.90	208.0
156.90	151.0
158.90	159.0
161.00	124.0
172.00	662.0
174.00	98880.0
175.00	7399.0
176.00	93328.0
177.00	6539.0
178.00	166.0

Instrument :
MSVOA_U
ClientSampleId :
BFB21

m/z	Abundance
35.90	1285.0
37.00	7283.0
38.00	6755.0
38.90	2857.0
39.90	447.0
44.00	1545.0
45.00	1313.0
45.80	126.0
47.10	1661.0
48.00	885.0
49.00	6928.0
50.00	31872.0
51.00	9866.0
52.00	481.0
54.90	259.0
56.00	1969.0
57.00	4106.0
57.90	112.0
60.00	1101.0
61.00	7875.0
62.00	7876.0
63.00	5441.0
64.00	561.0
65.00	381.0
67.10	482.0
68.00	16134.0
69.00	16832.0
70.00	1305.0
71.10	171.0
72.00	817.0
73.00	6950.0
74.00	26200.0
75.00	78792.0
76.00	6437.0
77.00	1074.0
78.00	763.0
78.90	4236.0
79.90	1368.0
80.90	4275.0
81.90	1075.0
83.00	121.0
86.00	239.0
87.00	6874.0
87.90	6233.0
90.80	476.0
92.00	4198.0
93.00	6541.0
94.00	16353.0
95.00	150720.0
96.00	10562.0
97.10	436.0
104.00	629.0
104.80	226.0
105.90	520.0
114.90	173.0
115.90	537.0
116.90	658.0
117.90	512.0
119.00	939.0
123.80	100.0
128.00	372.0
128.90	264.0
130.00	459.0
130.80	107.0
134.90	230.0
136.90	232.0
140.10	188.0
141.00	1166.0
141.90	126.0
143.00	1522.0
143.80	153.0
145.00	142.0
145.90	279.0
147.90	342.0
149.00	114.0
149.90	135.0
154.90	232.0
156.90	187.0
158.90	104.0
161.00	115.0
172.00	781.0
174.00	110376.0
175.00	8232.0
176.00	105176.0
177.00	6902.0
178.10	165.0

Instrument :
MSVOA_U
ClientSampleId :
BFB21

m/z	Abundance
36.00	1271.0
37.00	7336.0
38.00	6536.0
39.00	2803.0
39.90	381.0
44.00	1647.0
44.90	1198.0
45.90	125.0
47.10	1679.0
48.00	855.0
49.00	6716.0
50.00	31048.0
51.00	9755.0
52.00	353.0
55.00	307.0
55.90	2026.0
57.00	3978.0
60.00	1096.0
61.00	7702.0
62.00	7780.0
63.00	5442.0
64.00	559.0
65.00	424.0
67.20	513.0
68.00	15988.0
69.00	16496.0
70.00	1154.0
71.10	116.0
72.00	792.0
73.00	6816.0
74.00	24992.0
75.00	77280.0
76.00	6139.0
77.00	986.0
78.00	726.0
78.90	4165.0
79.90	1482.0
80.90	4317.0
81.90	980.0
86.00	220.0
87.00	6456.0
87.90	6028.0
90.80	575.0
92.00	4054.0
93.00	6302.0
94.00	15686.0
95.00	148160.0
96.00	10208.0
97.10	303.0
104.00	549.0
104.80	211.0
105.80	540.0
111.00	117.0
111.90	115.0
115.00	148.0
115.90	522.0
117.00	615.0
117.90	569.0
119.00	862.0
123.80	109.0
127.90	442.0
128.90	234.0
129.90	440.0
130.90	154.0
134.90	209.0
136.90	258.0
140.10	194.0
141.00	1270.0
142.00	171.0
143.00	1477.0
143.80	196.0
145.00	183.0
145.90	282.0
147.90	368.0
149.00	119.0
149.90	146.0
154.80	222.0
156.90	194.0
172.00	799.0
174.00	111928.0
175.00	8378.0
176.00	108288.0
177.00	6759.0
178.10	164.0

Instrument :
MSVOA_U
ClientSampleId :
BFB21