

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU051222\
 Data File : VU048531.D
 Acq On : 12 May 2022 11:00
 Operator : SY/MD
 Sample : VU0512MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK123

Quant Time: May 13 03:57:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM050622WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri May 13 03:54:05 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	261739	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	263421	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	138307	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	84621	37.540	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	75.080%
7) Chloroethane-d5	1.913	69	67077	45.595	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	91.200%
11) 1,1-Dichloroethene-d2	2.568	63	138591	34.546	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	69.100%
21) 2-Butanone-d5	4.620	46	176710	92.050	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	92.050%
24) Chloroform-d	5.064	84	203905	52.079	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.160%
26) 1,2-Dichloroethane-d4	5.703	65	149834	58.872	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	117.740%
32) Benzene-d6	5.729	84	380241	47.501	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.000%
36) 1,2-Dichloropropane-d6	6.690	67	117513	45.018	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	90.040%
41) Toluene-d8	7.899	98	343963	48.585	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.180%
43) trans-1,3-Dichloroprop...	8.179	79	65223	54.287	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	108.580%
47) 2-Hexanone-d5	8.629	63	124706	98.704	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	98.700%
56) 1,1,2,2-Tetrachloroeth...	10.755	84	183925	46.257	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	92.520%
66) 1,2-Dichlorobenzene-d4	12.195	152	144396	52.724	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	105.440%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU051222\
Data File : VU048531.D
Acq On : 12 May 2022 11:00
Operator : SY/MD
Sample : VU0512MBL01
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK123

Quant Time: May 13 03:57:13 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM050622WMA.M
Quant Title : VOC Analysis
QLast Update : Fri May 13 03:54:05 2022
Response via : Initial Calibration

