

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU051319\
 Data File : VU031959.D
 Acq On : 13 May 2019 09:55
 Operator : JC/SP
 Sample : VU0513WBL01
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK15

Quant Time: May 14 02:47:04 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR050819WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue May 14 02:38:02 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	412978	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	435019	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	171375	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	163077	5.80	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	116.00%
7) Chloroethane-d5	1.68	69	150526	6.48	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	129.60%
11) 1,1-Dichloroethene-d2	2.27	63	223219	3.68	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	73.60%
20) 2-Butanone-d5	4.21	46	552631	50.29	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	100.58%
24) Chloroform-d	4.64	84	303721	5.26	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.20%
26) 1,2-Dichloroethane-d4	5.31	65	168582	4.81	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.20%
32) Benzene-d6	5.33	84	565074	5.08	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.60%
36) 1,2-Dichloropropane-d6	6.33	67	201528	4.93	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.60%
41) Toluene-d8	7.56	98	417136	4.38	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.60%
43) trans-1,3-Dichloropropene-	7.85	79	64351	4.48	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.60%
46) 2-Hexanone-d5	8.31	63	393222	54.18	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	108.36%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	177818	5.58	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	111.60%
64) 1,2-Dichlorobenzene-d4	11.86	152	167382	5.91	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	118.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU051319\
Data File : VU031959.D
Acq On : 13 May 2019 09:55
Operator : JC/SP
Sample : VU0513WBL01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK15

Quant Time: May 14 02:47:04 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR050819WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue May 14 02:38:02 2019
Response via : Initial Calibration

