

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU051319\
 Data File : VU031961.D
 Acq On : 13 May 2019 10:50
 Operator : JC/SP
 Sample : K2791-14DL 5X
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F9D53DL

Quant Time: May 14 02:53:22 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR050819WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue May 14 02:38:02 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	473389	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	488349	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	181685	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	114562	3.55	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	71.00%
7) Chloroethane-d5	1.68	69	114387	4.30	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	86.00%
11) 1,1-Dichloroethene-d2	2.27	63	179567	2.58	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	51.60%#
20) 2-Butanone-d5	4.20	46	514596	40.85	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	81.70%
24) Chloroform-d	4.64	84	272863	4.12	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.40%
26) 1,2-Dichloroethane-d4	5.31	65	151206	3.76	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	75.20%
32) Benzene-d6	5.33	84	511761	4.10	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.00%
36) 1,2-Dichloropropane-d6	6.33	67	186332	4.06	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	81.20%
41) Toluene-d8	7.56	98	381424	3.57	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	71.40%
43) trans-1,3-Dichloropropene-	7.85	79	58463	3.63	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	72.60%
46) 2-Hexanone-d5	8.31	63	374952	46.02	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	92.04%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	164885	4.61	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.20%
64) 1,2-Dichlorobenzene-d4	11.86	152	145395	4.84	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.80%

Target Compounds

					Ovalue
13) Acetone	2.34	43	79365	9.225 ug/L	80
19) 1,1-Dichloroethane	3.44	63	31496	0.369 ug/L #	95
22) cis-1,2-Dichloroethene	4.22	96	342012	8.210 ug/L	80
27) 1,2-Dichloroethane	5.41	62	17423	0.314 ug/L	98
34) Trichloroethene	6.19	95	31670	0.738 ug/L	95
47) Tetrachloroethene	8.22	164	96631	3.462 ug/L	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU051319\
Data File : VU031961.D
Acq On : 13 May 2019 10:50
Operator : JC/SP
Sample : K2791-14DL 5X
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F9D53DL

Quant Time: May 14 02:53:22 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR050819WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue May 14 02:38:02 2019
Response via : Initial Calibration

