

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU051319\
 Data File : VU031969.D
 Acq On : 13 May 2019 14:08
 Operator : JC/SP
 Sample : VIBLK40
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK40

Quant Time: May 14 04:16:24 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR050819WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue May 14 02:38:02 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	476418	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	473889	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	187342	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	163195	5.03	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	100.60%
7) Chloroethane-d5	1.68	69	157161	5.87	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	117.40%
11) 1,1-Dichloroethene-d2	2.27	63	234454	3.35	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.00%
20) 2-Butanone-d5	4.19	46	598268	47.19	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	94.38%
24) Chloroform-d	4.64	84	324639	4.88	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.60%
26) 1,2-Dichloroethane-d4	5.31	65	183796	4.55	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.00%
32) Benzene-d6	0.00	84	0d	0.00	ug/L	
Spiked Amount	5.000	Range	70 - 125	Recovery	=	0.00%#
36) 1,2-Dichloropropane-d6	6.33	67	223665	5.02	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.40%
41) Toluene-d8	7.56	98	473621	4.56	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.20%
43) trans-1,3-Dichloropropene-	7.85	79	68937	4.41	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	88.20%
46) 2-Hexanone-d5	8.31	63	478744	60.55	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	121.10%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	188974	5.44	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	108.80%
64) 1,2-Dichlorobenzene-d4	11.85	152	192083	6.20	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	124.00%#

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU051319\
Data File : VU031969.D
Acq On : 13 May 2019 14:08
Operator : JC/SP
Sample : VIBLK40
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK40

Quant Time: May 14 04:16:24 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR050819WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue May 14 02:38:02 2019
Response via : Initial Calibration

