

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU051319\
 Data File : VU031976.D
 Acq On : 13 May 2019 18:17
 Operator : JC/SP
 Sample : K2744-17
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F9D19

Quant Time: May 14 04:46:49 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR050819WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue May 14 02:38:02 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	513053	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	519606	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	193100	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	116806	3.34	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	66.80%
7) Chloroethane-d5	1.67	69	122185	4.24	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	84.80%
11) 1,1-Dichloroethene-d2	2.27	63	185586	2.46	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	49.20%#
20) 2-Butanone-d5	4.18	46	508024	37.21	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	74.42%
24) Chloroform-d	4.64	84	272444	3.80	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	76.00%
26) 1,2-Dichloroethane-d4	5.31	65	158455	3.64	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	72.80%
32) Benzene-d6	5.33	84	537947	4.05	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.00%
36) 1,2-Dichloropropane-d6	6.33	67	187264	3.83	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	76.60%
41) Toluene-d8	7.56	98	406110	3.57	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	71.40%
43) trans-1,3-Dichloropropene-	7.85	79	49735	2.90	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	58.00%
46) 2-Hexanone-d5	8.31	63	364441	42.04	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	84.08%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	165045	4.34	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	86.80%
64) 1,2-Dichlorobenzene-d4	11.86	152	148717	4.66	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	93.20%

Target Compounds					Ovalue
14) Carbon disulfide	2.48	76	13014	0.106 ug/L #	83
30) Cyclohexane	4.99	56	70687	0.906 ug/L	87
47) Tetrachloroethene	8.23	164	4571	0.154 ug/L	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU051319\
Data File : VU031976.D
Acq On : 13 May 2019 18:17
Operator : JC/SP
Sample : K2744-17
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F9D19

Quant Time: May 14 04:46:49 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR050819WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue May 14 02:38:02 2019
Response via : Initial Calibration

