

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU051321\
 Data File : VU043708.D
 Acq On : 13 May 2021 16:22
 Operator : SY/MD
 Sample : VIBLK131
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK131

Quant Time: May 14 01:50:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR050321WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri May 14 01:48:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.256	114	90588	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.424	117	92154	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.819	152	39805	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	22422	3.08	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	61.60%
7) Chloroethane-d5	1.919	69	20629	3.87	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	77.40%
11) 1,1-Dichloroethene-d2	2.572	65	11544	3.23	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	64.60%
20) 2-Butanone-d5	4.642	46	118413	52.36	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.72%
24) Chloroform-d	5.073	84	59449	4.27	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.40%
26) 1,2-Dichloroethane-d4	5.710	65	40016	4.77	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.40%
32) Benzene-d6	5.736	84	100992	4.05	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.00%
36) 1,2-Dichloropropane-d6	6.700	67	40385	4.62	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.40%
41) Toluene-d8	7.906	98	80355	3.51	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	70.20%
43) trans-1,3-Dichloroprop...	8.186	79	14198	4.03	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	80.60%
46) 2-Hexanone-d5	8.642	63	92443	62.50	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	125.00%
56) 1,1,2,2-Tetrachloroeth...	10.764	84	36300	5.18	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	103.60%
66) 1,2-Dichlorobenzene-d4	12.202	152	32600	4.59	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	91.80%
Target Compounds						Qvalue
16) Methylene chloride	3.051	84	17104	2.11	ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU051321\
Data File : VU043708.D
Acq On : 13 May 2021 16:22
Operator : SY/MD
Sample : VIBLK131
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK131

Quant Time: May 14 01:50:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR050321WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri May 14 01:48:16 2021
Response via : Initial Calibration

