

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU051419\  
 Data File : VU031993.D  
 Acq On : 14 May 2019 14:41  
 Operator : JC/SP  
 Sample : K2741-06  
 Misc : 5.0mL/MSVOA U/WATER  
 ALS Vial : 13 Sample Multiplier: 1

Instrument :  
 MSVOA\_U  
 ClientSampleId :  
 VHBLK01

Quant Time: May 15 04:45:28 2019  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_U\METHOD\SOMULM051419WMA.M  
 Quant Title : VOC Analysis  
 QLast Update : Wed May 15 04:26:34 2019  
 Response via : Initial Calibration

| Internal Standards         | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|----------------------------|-------|------|----------|-------|-------|----------|
| 1) 1,4-Difluorobenzene     | 5.88  | 114  | 1069792  | 50.00 | ug/L  | 0.00     |
| 28) Chlorobenzene-d5       | 9.09  | 117  | 1051571  | 50.00 | ug/L  | 0.00     |
| 60) 1,4-Dichlorobenzene-d4 | 11.48 | 152  | 406428   | 50.00 | ug/L  | 0.00     |

## System Monitoring Compounds

|                                |         |       |          |          |      |         |
|--------------------------------|---------|-------|----------|----------|------|---------|
| 4) Vinyl Chloride-d3           | 1.40    | 65    | 430778   | 45.89    | ug/L | 0.00    |
| Spiked Amount                  | 50.000  | Range | 60 - 135 | Recovery | =    | 91.78%  |
| 7) Chloroethane-d5             | 1.67    | 69    | 370161   | 48.70    | ug/L | 0.00    |
| Spiked Amount                  | 50.000  | Range | 70 - 130 | Recovery | =    | 97.40%  |
| 11) 1,1-Dichloroethene-d2      | 2.27    | 63    | 609480   | 36.22    | ug/L | 0.00    |
| Spiked Amount                  | 50.000  | Range | 60 - 125 | Recovery | =    | 72.44%  |
| 21) 2-Butanone-d5              | 4.17    | 46    | 659022   | 100.25   | ug/L | 0.00    |
| Spiked Amount                  | 100.000 | Range | 40 - 130 | Recovery | =    | 100.25% |
| 24) Chloroform-d               | 4.64    | 84    | 729494   | 46.42    | ug/L | 0.00    |
| Spiked Amount                  | 50.000  | Range | 70 - 125 | Recovery | =    | 92.84%  |
| 26) 1,2-Dichloroethane-d4      | 5.31    | 65    | 476057   | 48.10    | ug/L | 0.00    |
| Spiked Amount                  | 50.000  | Range | 70 - 125 | Recovery | =    | 96.20%  |
| 32) Benzene-d6                 | 5.33    | 84    | 1568946  | 51.54    | ug/L | 0.00    |
| Spiked Amount                  | 50.000  | Range | 70 - 125 | Recovery | =    | 103.08% |
| 36) 1,2-Dichloropropane-d6     | 6.32    | 67    | 504413   | 50.22    | ug/L | 0.00    |
| Spiked Amount                  | 50.000  | Range | 70 - 120 | Recovery | =    | 100.44% |
| 41) Toluene-d8                 | 7.56    | 98    | 1346001  | 49.35    | ug/L | 0.00    |
| Spiked Amount                  | 50.000  | Range | 80 - 120 | Recovery | =    | 98.70%  |
| 43) trans-1,3-Dichloropropene- | 7.85    | 79    | 204313   | 46.11    | ug/L | 0.00    |
| Spiked Amount                  | 50.000  | Range | 60 - 125 | Recovery | =    | 92.22%  |
| 47) 2-Hexanone-d5              | 8.31    | 63    | 419573   | 103.01   | ug/L | 0.00    |
| Spiked Amount                  | 100.000 | Range | 45 - 130 | Recovery | =    | 103.01% |
| 57) 1,1,2,2-Tetrachloroethane- | 10.43   | 84    | 707867   | 47.61    | ug/L | 0.00    |
| Spiked Amount                  | 50.000  | Range | 65 - 120 | Recovery | =    | 95.22%  |
| 64) 1,2-Dichlorobenzene-d4     | 11.86   | 152   | 449554   | 52.57    | ug/L | 0.00    |
| Spiked Amount                  | 50.000  | Range | 80 - 120 | Recovery | =    | 105.14% |

| Target Compounds | Ovalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU051419\  
Data File : VU031993.D  
Acq On : 14 May 2019 14:41  
Operator : JC/SP  
Sample : K2741-06  
Misc : 5.0mL/MSVOA U/WATER  
ALS Vial : 13 Sample Multiplier: 1

Instrument :  
MSVOA\_U  
ClientSampleId :  
VHBLK01

Quant Time: May 15 04:45:28 2019  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_U\METHOD\SOMULM051419WMA.M  
Quant Title : VOC Analysis  
QLast Update : Wed May 15 04:26:34 2019  
Response via : Initial Calibration

