

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU051421\
 Data File : VU043720.D
 Acq On : 14 May 2021 00:02
 Operator : SY/MD
 Sample : VU0514WBL01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK150

Quant Time: May 14 04:43:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM051421WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri May 14 04:42:41 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.256	114	487911	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.423	117	462375	50.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.819	152	218313	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	196016	48.52	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	97.04%
7) Chloroethane-d5	1.909	69	141095	48.57	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	97.14%
11) 1,1-Dichloroethene-d2	2.571	63	249616	36.29	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	72.58%
21) 2-Butanone-d5	4.626	46	326380	103.28	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	103.28%
24) Chloroform-d	5.070	84	322256	47.01	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.02%
26) 1,2-Dichloroethane-d4	5.710	65	206874	48.68	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.36%
32) Benzene-d6	5.735	84	689854	50.34	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.68%
36) 1,2-Dichloropropane-d6	6.697	67	228959	51.23	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	102.46%
41) Toluene-d8	7.906	98	603334	48.93	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.86%
43) trans-1,3-Dichloroprop...	8.185	79	93813	48.01	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	96.02%
47) 2-Hexanone-d5	8.639	63	217611	101.68	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	101.68%
56) 1,1,2,2-Tetrachloroeth...	10.764	84	321866	47.59	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	95.18%
66) 1,2-Dichlorobenzene-d4	12.198	152	225627	52.37	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	104.74%
Target Compounds						
16) Methylene chloride	3.051	84	5568	1.06	ug/L	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU051421\
Data File : VU043720.D
Acq On : 14 May 2021 00:02
Operator : SY/MD
Sample : VU0514WBL01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK150

Quant Time: May 14 04:43:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM051421WMA.M
Quant Title : VOC Analysis
QLast Update : Fri May 14 04:42:41 2021
Response via : Initial Calibration

