

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU051520\
 Data File : VU038206.D
 Acq On : 15 May 2020 18:23
 Operator : JC/MD
 Sample : L2647-03
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK01

Quant Time: May 16 07:31:47 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR050720WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat May 16 03:30:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	323285	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	310558	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	144441	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	132955	4.50	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	90.00%
7) Chloroethane-d5	1.93	69	120008	4.69	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	93.80%
11) 1,1-Dichloroethene-d2	2.59	63	154168	3.12	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	62.40%
20) 2-Butanone-d5	4.67	46	445372	61.04	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	122.08%
24) Chloroform-d	5.10	84	225094	5.02	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.40%
26) 1,2-Dichloroethane-d4	5.74	65	135267	5.26	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.20%
32) Benzene-d6	5.76	84	459196	4.93	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.60%
36) 1,2-Dichloropropane-d6	6.72	67	157690	5.25	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	105.00%
41) Toluene-d8	7.92	98	382108	4.57	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.40%
43) trans-1,3-Dichloropropene-	8.20	79	56781	4.72	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	94.40%
46) 2-Hexanone-d5	8.65	63	329702	58.03	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	116.06%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	129604	5.47	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	109.40%
64) 1,2-Dichlorobenzene-d4	12.20	152	146433	5.21	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	104.20%

Target Compounds					Qvalue
16) Methylene chloride	3.08	84	3451	0.13 ug/L	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU051520\
Data File : VU038206.D
Acq On : 15 May 2020 18:23
Operator : JC/MD
Sample : L2647-03
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK01

Quant Time: May 16 07:31:47 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR050720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat May 16 03:30:00 2020
Response via : Initial Calibration

