

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU051521\
 Data File : VU043763.D
 Acq On : 14 May 2021 18:39
 Operator : SY/MD
 Sample : M2376-01DL 4X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BG1Y7DL

Quant Time: May 15 01:19:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR051521WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat May 15 00:49:19 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.256	114	226443	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.423	117	224853	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.819	152	105772	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	87443	5.24	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	104.80%
7) Chloroethane-d5	1.919	69	62787	5.38	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	107.60%
11) 1,1-Dichloroethene-d2	2.571	65	37456	5.41	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	108.20%
20) 2-Butanone-d5	4.645	46	162010	51.03	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	102.06%
24) Chloroform-d	5.073	84	141333	5.21	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.20%
26) 1,2-Dichloroethane-d4	5.710	65	74940	5.39	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.80%
32) Benzene-d6	5.735	84	303123	5.31	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	106.20%
36) 1,2-Dichloropropane-d6	6.700	67	95471	5.25	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	105.00%
41) Toluene-d8	7.906	98	256446	4.86	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.20%
43) trans-1,3-Dichloroprop...	8.185	79	33784	5.00	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	100.00%
46) 2-Hexanone-d5	8.642	63	154736	50.85	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	101.70%
56) 1,1,2,2-Tetrachloroeth...	10.764	84	74802	5.01	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.20%
66) 1,2-Dichlorobenzene-d4	12.198	152	96611	5.48	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	109.60%
Target Compounds						
					Qvalue	
5) Vinyl chloride	1.607	62	46582	2.42	ug/L	99
19) 1,1-Dichloroethane	3.880	63	124659	4.40	ug/L	99
22) cis-1,2-Dichloroethene	4.674	96	131601	7.98	ug/L	98
34) Trichloroethene	6.549	95	3255	0.20	ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU051521\
Data File : VU043763.D
Acq On : 14 May 2021 18:39
Operator : SY/MD
Sample : M2376-01DL 4X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BG1Y7DL

Quant Time: May 15 01:19:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR051521WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat May 15 00:49:19 2021
Response via : Initial Calibration

