

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU051619\
 Data File : VU032049.D
 Acq On : 16 May 2019 15:16
 Operator : JC/SP
 Sample : VU0516WBL01
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK16

Quant Time: May 17 07:33:44 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR051619WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri May 17 07:17:36 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	402715	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	415347	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	171166	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	166295	4.90	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	98.00%
7) Chloroethane-d5	1.68	69	146705	5.06	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	101.20%
11) 1,1-Dichloroethene-d2	2.27	63	225166	4.17	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	83.40%
20) 2-Butanone-d5	4.20	46	327077	49.11	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	98.22%
24) Chloroform-d	4.64	84	289167	4.85	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.00%
26) 1,2-Dichloroethane-d4	5.31	65	146817	4.77	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.40%
32) Benzene-d6	5.33	84	602409	4.75	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.00%
36) 1,2-Dichloropropane-d6	6.33	67	179240	4.84	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.80%
41) Toluene-d8	7.56	98	512830	4.59	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.80%
43) trans-1,3-Dichloropropene-	7.85	79	60353	4.53	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.60%
46) 2-Hexanone-d5	8.31	63	284294	47.22	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	94.44%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	140371	4.64	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.80%
64) 1,2-Dichlorobenzene-d4	11.86	152	181734	5.01	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU051619\
Data File : VU032049.D
Acq On : 16 May 2019 15:16
Operator : JC/SP
Sample : VU0516WBL01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK16

Quant Time: May 17 07:33:44 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR051619WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri May 17 07:17:36 2019
Response via : Initial Calibration

