

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU051619\
 Data File : VU032057.D
 Acq On : 16 May 2019 18:35
 Operator : JC/SP
 Sample : K2739-05
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK01

Quant Time: May 17 08:02:05 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR051619WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri May 17 07:17:36 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	403510	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	408835	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	159019	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	157666	4.64	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	92.80%
7) Chloroethane-d5	1.67	69	148251	5.10	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	102.00%
11) 1,1-Dichloroethene-d2	2.27	63	223357	4.12	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	82.40%
20) 2-Butanone-d5	4.19	46	326149	48.88	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	97.76%
24) Chloroform-d	4.64	84	289996	4.85	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.00%
26) 1,2-Dichloroethane-d4	5.31	65	150100	4.87	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.40%
32) Benzene-d6	5.33	84	609294	4.88	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.60%
36) 1,2-Dichloropropane-d6	6.33	67	178093	4.89	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.80%
41) Toluene-d8	7.56	98	505876	4.60	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.00%
43) trans-1,3-Dichloropropene-	7.85	79	56327	4.29	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.80%
46) 2-Hexanone-d5	8.31	63	294559	49.71	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	99.42%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	146576	4.92	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.40%
64) 1,2-Dichlorobenzene-d4	11.86	152	180432	5.35	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	107.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU051619\
Data File : VU032057.D
Acq On : 16 May 2019 18:35
Operator : JC/SP
Sample : K2739-05
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK01

Quant Time: May 17 08:02:05 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR051619WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri May 17 07:17:36 2019
Response via : Initial Calibration

