

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU051719\
 Data File : VU032141.D
 Acq On : 18 May 2019 08:44
 Operator : JC/SP
 Sample : K2860-16
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F9W49

Quant Time: May 20 05:21:56 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR051619WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon May 20 04:57:10 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	342976	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	352525	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	140263	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	107216	3.71	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	74.20%
7) Chloroethane-d5	1.68	69	95645	3.87	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	77.40%
11) 1,1-Dichloroethene-d2	2.27	63	150803	3.28	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.60%
20) 2-Butanone-d5	4.19	46	225010	39.67	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	79.34%
24) Chloroform-d	4.64	84	212958	4.19	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.80%
26) 1,2-Dichloroethane-d4	5.31	65	108347	4.13	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.60%
32) Benzene-d6	5.33	84	429441	3.99	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	79.80%
36) 1,2-Dichloropropane-d6	6.33	67	126591	4.03	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	80.60%
41) Toluene-d8	7.56	98	350128	3.70	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	74.00%
43) trans-1,3-Dichloropropene-	7.85	79	38247	3.38	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	67.60%
46) 2-Hexanone-d5	8.31	63	197695	38.69	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	77.38%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	106423	4.14	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	82.80%
64) 1,2-Dichlorobenzene-d4	11.86	152	136998	4.61	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	92.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU051719\
Data File : VU032141.D
Acq On : 18 May 2019 08:44
Operator : JC/SP
Sample : K2860-16
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F9W49

Quant Time: May 20 05:21:56 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR051619WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon May 20 04:57:10 2019
Response via : Initial Calibration

