

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU051721\
 Data File : VU043782.D
 Acq On : 17 May 2021 10:30
 Operator : SY/MD
 Sample : VU0517WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK151

Quant Time: May 17 13:16:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR051521WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat May 15 00:49:19 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.260	114	226955	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.424	117	228071	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.819	152	107666	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	83114	4.97	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	99.40%
7) Chloroethane-d5	1.919	69	60419	5.17	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	103.40%
11) 1,1-Dichloroethene-d2	2.575	65	35302	5.09	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	101.80%
20) 2-Butanone-d5	4.649	46	147436	46.34	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	92.68%
24) Chloroform-d	5.073	84	139931	5.14	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.80%
26) 1,2-Dichloroethane-d4	5.713	65	71879	5.16	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.20%
32) Benzene-d6	5.736	84	296537	5.12	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.40%
36) 1,2-Dichloropropane-d6	6.700	67	92803	5.03	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.60%
41) Toluene-d8	7.906	98	252706	4.72	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.40%
43) trans-1,3-Dichloroprop...	8.189	79	34416	5.02	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	100.40%
46) 2-Hexanone-d5	8.645	63	152440	49.39	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	98.78%
56) 1,1,2,2-Tetrachloroeth...	10.764	84	76024	5.02	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.40%
66) 1,2-Dichlorobenzene-d4	12.202	152	96588	5.38	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	107.60%
Target Compounds						
					Qvalue	
16) Methylene chloride	3.054	84	3003	0.14	ug/L	97
70) 1,2,4-trichlorobenzene	13.854	180	2186	0.11	ug/L	96
72) 1,2,3-Trichlorobenzene	14.337	180	2104	0.11	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU051721\
Data File : VU043782.D
Acq On : 17 May 2021 10:30
Operator : SY/MD
Sample : VU0517WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK151

Quant Time: May 17 13:16:28 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR051521WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat May 15 00:49:19 2021
Response via : Initial Calibration

