

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU051721\
 Data File : VU043783.D
 Acq On : 17 May 2021 10:59
 Operator : SY/MD
 Sample : M2418-08
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BG201

Quant Time: May 17 11:45:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR051521WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat May 15 00:49:19 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.260	114	233814	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.427	117	235057	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.819	152	111205	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	82310	4.78	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	95.60%
7) Chloroethane-d5	1.919	69	59735	4.96	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	99.20%
11) 1,1-Dichloroethene-d2	2.575	65	35312	4.94	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	98.80%
20) 2-Butanone-d5	4.645	46	185533	56.60	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	113.20%
24) Chloroform-d	5.073	84	138909	4.96	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.20%
26) 1,2-Dichloroethane-d4	5.713	65	71836	5.00	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.00%
32) Benzene-d6	5.739	84	293981	4.92	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.40%
36) 1,2-Dichloropropane-d6	6.700	67	93413	4.91	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.20%
41) Toluene-d8	7.906	98	250119	4.54	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.80%
43) trans-1,3-Dichloroprop...	8.189	79	33528	4.75	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	95.00%
46) 2-Hexanone-d5	8.645	63	154102	48.44	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	96.88%
56) 1,1,2,2-Tetrachloroeth...	10.764	84	74988	4.80	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.00%
66) 1,2-Dichlorobenzene-d4	12.201	152	98573	5.32	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	106.40%
Target Compounds						
					Qvalue	
13) Acetone	2.655	43	17989	8.64	ug/L	95
21) 2-Butanone	4.729	43	8194	2.06	ug/L	92
42) Toluene	7.976	91	7377	0.10	ug/L	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU051721\
Data File : VU043783.D
Acq On : 17 May 2021 10:59
Operator : SY/MD
Sample : M2418-08
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BG201

Quant Time: May 17 11:45:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR051521WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat May 15 00:49:19 2021
Response via : Initial Calibration

