

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU051721\
 Data File : VU043796.D
 Acq On : 17 May 2021 17:00
 Operator : SY/MD
 Sample : M2380-07
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: May 18 03:12:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR051521WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat May 15 00:49:19 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	6.256	114	226466	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.423	117	229406	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.819	152	105603	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	76691	4.59	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	91.80%
7) Chloroethane-d5	1.919	69	57106	4.90	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	98.00%
11) 1,1-Dichloroethene-d2	2.571	65	33640	4.86	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	97.20%
20) 2-Butanone-d5	4.642	46	147956	46.60	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	93.20%
24) Chloroform-d	5.073	84	134680	4.96	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.20%
26) 1,2-Dichloroethane-d4	5.713	65	70273	5.05	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.00%
32) Benzene-d6	5.735	84	287095	4.93	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.60%
36) 1,2-Dichloropropane-d6	6.700	67	90048	4.85	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.00%
41) Toluene-d8	7.906	98	240929	4.48	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.60%
43) trans-1,3-Dichloroprop...	8.185	79	31395	4.55	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.00%
46) 2-Hexanone-d5	8.645	63	146160	47.08	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	94.16%
56) 1,1,2,2-Tetrachloroeth...	10.764	84	71641	4.70	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.00%
66) 1,2-Dichlorobenzene-d4	12.201	152	91611	5.21	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	104.20%
Target Compounds						
16) Methylene chloride	3.054	84	5978	0.27	ug/L	Qvalue 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU051721\
Data File : VU043796.D
Acq On : 17 May 2021 17:00
Operator : SY/MD
Sample : M2380-07
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: May 18 03:12:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR051521WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat May 15 00:49:19 2021
Response via : Initial Calibration

