

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU051722\
 Data File : VU048598.D
 Acq On : 17 May 2022 12:15
 Operator : SY/MD
 Sample : VU0517MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Quant Time: May 18 05:18:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM051622WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed May 18 05:17:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	216566	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	218155	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	118567	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	56231	42.392	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	84.780%
7) Chloroethane-d5	1.912	69	46306	46.597	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	93.200%
11) 1,1-Dichloroethene-d2	2.568	63	102650	34.218	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	68.440%
21) 2-Butanone-d5	4.629	46	122151	96.770	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	96.770%
24) Chloroform-d	5.066	84	160304	47.824	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.640%
26) 1,2-Dichloroethane-d4	5.703	65	118941	50.298	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.600%
32) Benzene-d6	5.729	84	278789	45.918	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.840%
36) 1,2-Dichloropropane-d6	6.690	67	81962	45.984	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	91.960%
41) Toluene-d8	7.899	98	258923	45.339	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.680%
43) trans-1,3-Dichloroprop...	8.179	79	48586	42.628	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.260%
47) 2-Hexanone-d5	8.635	63	87642	89.661	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	89.660%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	132021	44.763	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.520%
66) 1,2-Dichlorobenzene-d4	12.192	152	116651	48.126	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.260%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU051722\
Data File : VU048598.D
Acq On : 17 May 2022 12:15
Operator : SY/MD
Sample : VU0517MBL01
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 4 Sample Multiplier: 1

Quant Time: May 18 05:18:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM051622WMA.M
Quant Title : VOC Analysis
QLast Update : Wed May 18 05:17:36 2022
Response via : Initial Calibration

