

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU051722\
 Data File : VU048603.D
 Acq On : 17 May 2022 16:45
 Operator : SY/MD
 Sample : N2849-07
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Quant Time: May 18 05:19:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM051622WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed May 18 05:17:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	234789	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	239505	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.816	152	138218	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	61505	42.769	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	85.540%
7) Chloroethane-d5	1.912	69	48018	44.569	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	89.140%
11) 1,1-Dichloroethene-d2	2.568	63	109511	33.672	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	67.340%
21) 2-Butanone-d5	4.620	46	123043	89.911	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	89.910%
24) Chloroform-d	5.067	84	170049	46.794	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.580%
26) 1,2-Dichloroethane-d4	5.703	65	124531	48.574	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.140%
32) Benzene-d6	5.729	84	303751	45.570	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.140%
36) 1,2-Dichloropropane-d6	6.694	67	89195	45.581	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	91.160%
41) Toluene-d8	7.899	98	288615	46.033	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.060%
43) trans-1,3-Dichloroprop...	8.182	79	52948	42.314	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	84.620%
47) 2-Hexanone-d5	8.632	63	91002	84.799	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	84.800%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	142362	43.966	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	87.940%
66) 1,2-Dichlorobenzene-d4	12.195	152	135953	48.115	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.240%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU051722\
Data File : VU048603.D
Acq On : 17 May 2022 16:45
Operator : SY/MD
Sample : N2849-07
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Quant Time: May 18 05:19:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM051622WMA.M
Quant Title : VOC Analysis
QLast Update : Wed May 18 05:17:36 2022
Response via : Initial Calibration

