

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU051821\
 Data File : VU043809.D
 Acq On : 18 May 2021 13:13
 Operator : SY/MD
 Sample : M2396-05DL 50X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0GJ0DL

Quant Time: May 19 01:41:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR051521WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed May 19 01:40:07 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.256	114	220114	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.423	117	227762	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.819	152	103877	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	72316	4.46	ug/L	0.00
Spiked Amount 5.000	Range 40 - 130		Recovery = 89.20%			
7) Chloroethane-d5	1.919	69	54097	4.77	ug/L	0.00
Spiked Amount 5.000	Range 65 - 130		Recovery = 95.40%			
11) 1,1-Dichloroethene-d2	2.572	65	32544	4.83	ug/L	0.00
Spiked Amount 5.000	Range 60 - 125		Recovery = 96.60%			
20) 2-Butanone-d5	4.645	46	156632	50.76	ug/L	0.00
Spiked Amount 50.000	Range 40 - 130		Recovery = 101.52%			
24) Chloroform-d	5.073	84	133108	5.04	ug/L	0.00
Spiked Amount 5.000	Range 70 - 125		Recovery = 100.80%			
26) 1,2-Dichloroethane-d4	5.713	65	68658	5.08	ug/L	0.00
Spiked Amount 5.000	Range 70 - 130		Recovery = 101.60%			
32) Benzene-d6	5.735	84	273009	4.72	ug/L	0.00
Spiked Amount 5.000	Range 70 - 125		Recovery = 94.40%			
36) 1,2-Dichloropropane-d6	6.700	67	90114	4.89	ug/L	0.00
Spiked Amount 5.000	Range 60 - 140		Recovery = 97.80%			
41) Toluene-d8	7.906	98	237369	4.44	ug/L	0.00
Spiked Amount 5.000	Range 70 - 130		Recovery = 88.80%			
43) trans-1,3-Dichloroprop...	8.186	79	30316	4.43	ug/L	0.00
Spiked Amount 5.000	Range 55 - 130		Recovery = 88.60%			
46) 2-Hexanone-d5	8.645	63	148140	48.06	ug/L	0.00
Spiked Amount 50.000	Range 45 - 130		Recovery = 96.12%			
56) 1,1,2,2-Tetrachloroeth...	10.764	84	71905	4.75	ug/L	0.00
Spiked Amount 5.000	Range 65 - 120		Recovery = 95.00%			
66) 1,2-Dichlorobenzene-d4	12.201	152	90471	5.23	ug/L	0.00
Spiked Amount 5.000	Range 80 - 120		Recovery = 104.60%			
Target Compounds						Qvalue
22) cis-1,2-Dichloroethene	4.674	96	8656	0.54	ug/L	97
31) Carbon tetrachloride	5.533	117	2122	0.11	ug/L	98
34) Trichloroethene	6.549	95	913824	54.06	ug/L	97
47) Tetrachloroethene	8.562	164	165891	12.14	ug/L	98
50) 1,2-Dibromoethane	8.931	107	11966	1.03	ug/L #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU051821\
Data File : VU043809.D
Acq On : 18 May 2021 13:13
Operator : SY/MD
Sample : M2396-05DL 50X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0GJ0DL

Quant Time: May 19 01:41:54 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR051521WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed May 19 01:40:07 2021
Response via : Initial Calibration

