

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU051821\
 Data File : VU043816.D
 Acq On : 18 May 2021 16:17
 Operator : SY/MD
 Sample : M2396-02
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: May 19 01:42:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR051521WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed May 19 01:40:07 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	6.259	114	225154	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.423	117	224849	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.819	152	106548	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	73424	4.42	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	88.40%
7) Chloroethane-d5	1.919	69	52942	4.57	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	91.40%
11) 1,1-Dichloroethene-d2	2.571	65	32815	4.76	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	95.20%
20) 2-Butanone-d5	4.645	46	149789	47.45	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	94.90%
24) Chloroform-d	5.073	84	131133	4.86	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.20%
26) 1,2-Dichloroethane-d4	5.713	65	68126	4.92	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.40%
32) Benzene-d6	5.735	84	276196	4.84	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.80%
36) 1,2-Dichloropropane-d6	6.700	67	87837	4.83	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.60%
41) Toluene-d8	7.906	98	233973	4.44	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.80%
43) trans-1,3-Dichloroprop...	8.185	79	30515	4.52	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.40%
46) 2-Hexanone-d5	8.645	63	143272	47.08	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	94.16%
56) 1,1,2,2-Tetrachloroeth...	10.764	84	70103	4.70	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.00%
66) 1,2-Dichlorobenzene-d4	12.201	152	90990	5.13	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.60%
Target Compounds						
16) Methylene chloride	3.054	84	9920	0.45	ug/L	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU051821\
Data File : VU043816.D
Acq On : 18 May 2021 16:17
Operator : SY/MD
Sample : M2396-02
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: May 19 01:42:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR051521WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed May 19 01:40:07 2021
Response via : Initial Calibration

