

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU051822\
 Data File : VU048607.D
 Acq On : 18 May 2022 11:50
 Operator : SY/MD
 Sample : VU0518MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK130

Quant Time: May 19 06:02:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM051622WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu May 19 06:01:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	218788	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	222979	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	125714	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	54748	40.855	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	81.700%
7) Chloroethane-d5	1.912	69	44544	44.368	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	88.740%
11) 1,1-Dichloroethene-d2	2.568	63	104417	34.454	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	68.900%
21) 2-Butanone-d5	4.623	46	118371	92.823	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	92.820%
24) Chloroform-d	5.067	84	161129	47.582	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.160%
26) 1,2-Dichloroethane-d4	5.703	65	118656	49.667	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.340%
32) Benzene-d6	5.729	84	280692	45.232	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.460%
36) 1,2-Dichloropropane-d6	6.693	67	82852	45.478	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	90.960%
41) Toluene-d8	7.899	98	264551	45.322	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.640%
43) trans-1,3-Dichloroprop...	8.182	79	48625	41.740	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	83.480%
47) 2-Hexanone-d5	8.636	63	86507	86.585	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	86.590%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	132257	43.872	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	87.740%
66) 1,2-Dichlorobenzene-d4	12.195	152	118166	45.980	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.960%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU051822\
Data File : VU048607.D
Acq On : 18 May 2022 11:50
Operator : SY/MD
Sample : VU0518MBL01
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK130

Quant Time: May 19 06:02:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM051622WMA.M
Quant Title : VOC Analysis
QLast Update : Thu May 19 06:01:59 2022
Response via : Initial Calibration

