

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU051823\
 Data File : VU054176.D
 Acq On : 18 May 2023 19:41
 Operator : JC/MD
 Sample : 02844-10
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_U
ClientSampleId :
 GBX04

Quant Time: May 19 02:09:57 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR051823WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri May 19 02:02:28 2023
 Response via : Initial Calibration

Manual Integrations
APPROVED
 Reviewed By :Krupa Patel 05/19/2023
 Supervised By :Mahesh Dadoda 05/19/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	6.243	114	165554	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	159833	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	78529	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	45110	5.054	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	101.000%	
7) Chloroethane-d5	1.912	69	36323	5.440	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	108.800%	
11) 1,1-Dichloroethene-d2	2.565	65	20926	5.191	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	103.800%	
20) 2-Butanone-d5	4.632	46	87836	55.183	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	110.360%	
24) Chloroform-d	5.057	84	84312	5.024	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	100.400%	
26) 1,2-Dichloroethane-d4	5.697	65	41240	5.342	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	106.800%	
32) Benzene-d6	5.722	84	170780	5.247	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	105.000%	
36) 1,2-Dichloropropane-d6	6.684	67	51713	5.497	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	110.000%	
41) Toluene-d8	7.893	98	151019	4.913	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	98.200%	
43) trans-1,3-Dichloroprop...	8.176	79	18927	4.981	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	99.600%	
46) 2-Hexanone-d5	8.629	63	65046	57.491	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	114.980%	
56) 1,1,2,2-Tetrachloroeth...	10.751	84	36112	5.355	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	107.200%	
66) 1,2-Dichlorobenzene-d4	12.188	152	56478	5.894	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	117.800%	
Target Compounds						
13) Acetone	2.642	43	46150	21.313	ug/L	93
21) 2-Butanone	4.722	43	6623m	2.309	ug/L	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU051823\
Data File : VU054176.D
Acq On : 18 May 2023 19:41
Operator : JC/MD
Sample : 02844-10
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GBX04

Quant Time: May 19 02:09:57 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR051823WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri May 19 02:02:28 2023
Response via : Initial Calibration

Manual Integrations
APPROVED
Reviewed By :Krupa Patel 05/19/2023
Supervised By :Mahesh Dadoda 05/19/2023

