

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU051823\
 Data File : VU054194.D
 Acq On : 19 May 2023 03:44
 Operator : JC/MD
 Sample : 02843-03
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GBWX9

Quant Time: May 19 06:20:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR051823WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri May 19 04:28:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.244	114	183371	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	176925	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.806	152	87345	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	42952	4.345	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	86.800%
7) Chloroethane-d5	1.913	69	32944	4.454	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.000%
11) 1,1-Dichloroethene-d2	2.565	65	18500	4.143	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	82.800%
20) 2-Butanone-d5	4.633	46	82078	46.555	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	93.120%
24) Chloroform-d	5.057	84	79748	4.291	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.800%
26) 1,2-Dichloroethane-d4	5.697	65	38176	4.464	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.200%
32) Benzene-d6	5.723	84	160970	4.468	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.400%
36) 1,2-Dichloropropane-d6	6.684	67	48573	4.664	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.200%
41) Toluene-d8	7.893	98	141158	4.149	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.000%
43) trans-1,3-Dichloroprop...	8.176	79	17600	4.184	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.600%
46) 2-Hexanone-d5	8.629	63	59185	47.258	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	94.520%
56) 1,1,2,2-Tetrachloroeth...	10.751	84	34426	4.612	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.200%
66) 1,2-Dichlorobenzene-d4	12.189	152	52393	4.916	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.400%
Target Compounds						
					Qvalue	
13) Acetone	2.642	43	11791	4.916	ug/L	93
14) Carbon disulfide	2.790	76	9488	0.336	ug/L #	95
34) Trichloroethene	6.536	95	19012	1.412	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU051823\
Data File : VU054194.D
Acq On : 19 May 2023 03:44
Operator : JC/MD
Sample : 02843-03
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GBWX9

Quant Time: May 19 06:20:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR051823WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri May 19 04:28:00 2023
Response via : Initial Calibration

