

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU051823\
 Data File : VU054187.D
 Acq On : 19 May 2023 00:54
 Operator : JC/MD
 Sample : VU0518WBL02
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK193

Quant Time: May 19 04:28:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR051823WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri May 19 04:28:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.244	114	172645	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.411	117	160991	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.806	152	75851	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.598	65	46014	4.944	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	98.800%
7) Chloroethane-d5	1.913	69	35538	5.104	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	102.000%
11) 1,1-Dichloroethene-d2	2.562	65	21544	5.125	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	102.400%
20) 2-Butanone-d5	4.630	46	89343	53.824	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	107.640%
24) Chloroform-d	5.057	84	84367	4.821	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.400%
26) 1,2-Dichloroethane-d4	5.697	65	41763	5.187	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.800%
32) Benzene-d6	5.723	84	169232	5.162	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.200%
36) 1,2-Dichloropropane-d6	6.684	67	51344	5.418	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	108.400%
41) Toluene-d8	7.893	98	149498	4.829	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.600%
43) trans-1,3-Dichloroprop...	8.176	79	18817	4.916	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	98.400%
46) 2-Hexanone-d5	8.629	63	63177	55.438	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	110.880%
56) 1,1,2,2-Tetrachloroeth...	10.752	84	36019	5.303	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	106.000%
66) 1,2-Dichlorobenzene-d4	12.189	152	52244	5.644	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	112.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU051823\
Data File : VU054187.D
Acq On : 19 May 2023 00:54
Operator : JC/MD
Sample : VU0518WBL02
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK193

Quant Time: May 19 04:28:37 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR051823WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri May 19 04:28:00 2023
Response via : Initial Calibration

