

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU051823\
 Data File : VU054198.D
 Acq On : 19 May 2023 05:21
 Operator : JC/MD
 Sample : 02843-07
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GBWY3

Quant Time: May 19 06:21:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR051823WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri May 19 04:28:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.243	114	190704	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	181315	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.806	152	88758	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	42919	4.174	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	83.400%
7) Chloroethane-d5	1.912	69	34982	4.548	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	91.000%
11) 1,1-Dichloroethene-d2	2.565	65	20108	4.330	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	86.600%
20) 2-Butanone-d5	4.632	46	83717	45.659	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	91.320%
24) Chloroform-d	5.057	84	82008	4.242	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.800%
26) 1,2-Dichloroethane-d4	5.697	65	40971	4.607	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.200%
32) Benzene-d6	5.722	84	163653	4.432	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.600%
36) 1,2-Dichloropropane-d6	6.684	67	49261	4.616	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.400%
41) Toluene-d8	7.893	98	143843	4.126	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.600%
43) trans-1,3-Dichloroprop...	8.176	79	17337	4.022	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	80.400%
46) 2-Hexanone-d5	8.629	63	59900	46.671	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.340%
56) 1,1,2,2-Tetrachloroeth...	10.751	84	35241	4.607	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.200%
66) 1,2-Dichlorobenzene-d4	12.188	152	50624	4.674	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	93.400%
Target Compounds						Qvalue
34) Trichloroethene	6.536	95	5022	0.364	ug/L	89
47) Tetrachloroethene	8.552	164	2218	0.196	ug/L	90


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU051823\
Data File : VU054198.D
Acq On : 19 May 2023 05:21
Operator : JC/MD
Sample : 02843-07
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GBWY3

Quant Time: May 19 06:21:11 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR051823WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri May 19 04:28:00 2023
Response via : Initial Calibration









#34
Trichloroethene
Concen: 0.364 ug/L
RT: 6.536 min Scan# 1
Delta R.T. -0.000 min
Lab File: VU054198.D
Acq: 19 May 2023 05:21

Instrument :
MSVOA_U
ClientSampleId :
GBWY3

Tgt Ion: 95 Resp: 5022
Ion Ratio Lower Upper
95 100
97 68.2 45.8 85.0
132 88.3 70.1 130.3
130 88.9 72.4 134.5



#47
Tetrachloroethene
Concen: 0.196 ug/L
RT: 8.552 min Scan# 2243
Delta R.T. 0.003 min
Lab File: VU054198.D
Acq: 19 May 2023 05:21

Tgt Ion:164 Resp: 2218
Ion Ratio Lower Upper
164 100
129 105.5 63.3 117.7
131 94.2 61.0 113.4
166 134.6 88.6 164.6

