

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU051922\
 Data File : VU048624.D
 Acq On : 19 May 2022 10:40
 Operator : SY/MD
 Sample : VU0519MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK132

Quant Time: May 20 04:39:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM051622WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri May 20 04:37:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	213796	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	213582	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	120118	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	52961	40.444	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	80.880%
7) Chloroethane-d5	1.912	69	44720	45.584	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	91.160%
11) 1,1-Dichloroethene-d2	2.568	63	102802	34.713	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	69.420%
21) 2-Butanone-d5	4.623	46	121589	97.572	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	97.570%
24) Chloroform-d	5.063	84	159443	48.184	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.360%
26) 1,2-Dichloroethane-d4	5.703	65	117926	50.514	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.020%
32) Benzene-d6	5.729	84	281314	47.326	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.660%
36) 1,2-Dichloropropane-d6	6.690	67	82973	47.548	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	95.100%
41) Toluene-d8	7.899	98	264541	47.315	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.620%
43) trans-1,3-Dichloroprop...	8.179	79	51036	45.737	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	91.480%
47) 2-Hexanone-d5	8.632	63	89452	93.472	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	93.470%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	133311	46.168	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	92.340%
66) 1,2-Dichlorobenzene-d4	12.192	152	121239	49.373	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.740%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU051922\
Data File : VU048624.D
Acq On : 19 May 2022 10:40
Operator : SY/MD
Sample : VU0519MBL01
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK132

Quant Time: May 20 04:39:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM051622WMA.M
Quant Title : VOC Analysis
QLast Update : Fri May 20 04:37:51 2022
Response via : Initial Calibration

