

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU051922\
 Data File : VU048631.D
 Acq On : 19 May 2022 15:43
 Operator : SY/MD
 Sample : N2843-15ME
 Misc : 6.37g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EW4J5ME

Quant Time: May 20 04:41:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM051622WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri May 20 04:37:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	231084	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	233708	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	129871	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	50672	35.801	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	71.600%
7) Chloroethane-d5	1.912	69	42533	40.111	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	80.220%
11) 1,1-Dichloroethene-d2	2.568	63	98576	30.796	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	61.600%
21) 2-Butanone-d5	4.626	46	121807	90.435	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	90.430%
24) Chloroform-d	5.063	84	158581	44.338	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.680%
26) 1,2-Dichloroethane-d4	5.703	65	116467	46.157	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.320%
32) Benzene-d6	5.729	84	280462	43.120	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.240%
36) 1,2-Dichloropropane-d6	6.694	67	82648	43.283	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	86.560%
41) Toluene-d8	7.899	98	262996	42.988	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	85.980%
43) trans-1,3-Dichloroprop...	8.179	79	50463	41.329	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	82.660%
47) 2-Hexanone-d5	8.636	63	88917	84.912	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	84.910%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	133765	42.336	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	84.680%
66) 1,2-Dichlorobenzene-d4	12.195	152	121888	45.910	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.820%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU051922\
Data File : VU048631.D
Acq On : 19 May 2022 15:43
Operator : SY/MD
Sample : N2843-15ME
Misc : 6.37g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EW4J5ME

Quant Time: May 20 04:41:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM051622WMA.M
Quant Title : VOC Analysis
QLast Update : Fri May 20 04:37:51 2022
Response via : Initial Calibration

