

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU052019\
 Data File : VU032149.D
 Acq On : 20 May 2019 12:48
 Operator : JC/SP
 Sample : K2860-04DL 4X
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F9W31DL

Quant Time: May 21 07:55:25 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR051619WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue May 21 06:39:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	381122	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	387540	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	157626	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	106599	3.32	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	66.40%
7) Chloroethane-d5	1.68	69	96499	3.52	ug/L	0.01
Spiked Amount	5.000	Range	65 - 130	Recovery	=	70.40%
11) 1,1-Dichloroethene-d2	2.27	63	152640	2.98	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	59.60%#
20) 2-Butanone-d5	4.20	46	252730	40.10	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	80.20%
24) Chloroform-d	4.64	84	219135	3.88	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	77.60%
26) 1,2-Dichloroethane-d4	5.31	65	110861	3.81	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	76.20%
32) Benzene-d6	5.33	84	444395	3.76	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	75.20%
36) 1,2-Dichloropropane-d6	6.32	67	131187	3.80	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	76.00%
41) Toluene-d8	7.56	98	358513	3.44	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	68.80%#
43) trans-1,3-Dichloropropene-	7.85	79	46429	3.73	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	74.60%
46) 2-Hexanone-d5	8.31	63	209848	37.36	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	74.72%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	114615	4.06	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	81.20%
64) 1,2-Dichlorobenzene-d4	11.86	152	141024	4.22	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	84.40%

Target Compounds

					Ovalue
5) Vinyl chloride	1.40	62	2783	0.145 ug/L #	17
13) Acetone	2.34	43	27725	13.399 ug/L	95
18) trans-1,2-Dichloroethene	2.97	96	6702	0.409 ug/L	90
22) cis-1,2-Dichloroethene	4.22	96	44439	2.757 ug/L	85
34) Trichloroethene	6.18	95	80144	4.965 ug/L	94
47) Tetrachloroethene	8.23	164	1776	0.147 ug/L #	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU052019\
Data File : VU032149.D
Acq On : 20 May 2019 12:48
Operator : JC/SP
Sample : K2860-04DL 4X
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F9W31DL

Quant Time: May 21 07:55:25 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR051619WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue May 21 06:39:31 2019
Response via : Initial Calibration

