

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052021\
 Data File : VU043833.D
 Acq On : 19 May 2021 17:49
 Operator : SY/MD
 Sample : M2460-02
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AR5

Quant Time: May 20 03:56:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR051521WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu May 20 03:54:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.256	114	217537	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.423	117	219903	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.819	152	100359	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	65077	4.06	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	81.20%
7) Chloroethane-d5	1.916	69	45152	4.03	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	80.60%
11) 1,1-Dichloroethene-d2	2.572	65	29136	4.38	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	87.60%
20) 2-Butanone-d5	4.649	46	191403	62.76	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	125.52%
24) Chloroform-d	5.073	84	128890	4.94	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.80%
26) 1,2-Dichloroethane-d4	5.713	65	67208	5.03	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.60%
32) Benzene-d6	5.735	84	264732	4.74	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.80%
36) 1,2-Dichloropropane-d6	6.700	67	84972	4.77	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.40%
41) Toluene-d8	7.906	98	219091	4.25	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.00%
43) trans-1,3-Dichloroprop...	8.185	79	29043	4.40	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	88.00%
46) 2-Hexanone-d5	8.645	63	146341	49.17	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	98.34%
56) 1,1,2,2-Tetrachloroeth...	10.764	84	70474	4.83	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.60%
66) 1,2-Dichlorobenzene-d4	12.201	152	92687	5.54	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	110.80%
Target Compounds						Qvalue
3) Chloromethane	1.520	50	3060	0.16	ug/L	100
14) Carbon disulfide	2.800	76	6180	0.13	ug/L	# 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052021\
Data File : VU043833.D
Acq On : 19 May 2021 17:49
Operator : SY/MD
Sample : M2460-02
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AR5

Quant Time: May 20 03:56:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR051521WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu May 20 03:54:53 2021
Response via : Initial Calibration

