

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052021\
 Data File : VU043839.D
 Acq On : 19 May 2021 20:14
 Operator : SY/MD
 Sample : M2464-05
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 GB3Y8

Quant Time: May 20 03:57:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR051521WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu May 20 03:54:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	6.259	114	218365	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.423	117	220720	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.819	152	101623	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	65488	4.07	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	81.40%	
7) Chloroethane-d5	1.916	69	44201	3.93	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	78.60%	
11) 1,1-Dichloroethene-d2	2.571	65	28905	4.33	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	86.60%	
20) 2-Butanone-d5	4.642	46	191329	62.50	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	125.00%	
24) Chloroform-d	5.073	84	128671	4.92	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	98.40%	
26) 1,2-Dichloroethane-d4	5.710	65	66799	4.98	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	99.60%	
32) Benzene-d6	5.735	84	265379	4.73	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	94.60%	
36) 1,2-Dichloropropane-d6	6.700	67	85677	4.80	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	96.00%	
41) Toluene-d8	7.906	98	219660	4.24	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	84.80%	
43) trans-1,3-Dichloroprop...	8.189	79	28682	4.33	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	86.60%	
46) 2-Hexanone-d5	8.642	63	152281	50.98	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	101.96%	
56) 1,1,2,2-Tetrachloroeth...	10.764	84	71809	4.90	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	98.00%	
66) 1,2-Dichlorobenzene-d4	12.201	152	89571	5.29	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	105.80%	
Target Compounds						
13) Acetone	2.649	43	5951	3.06	ug/L	94
42) Toluene	7.980	91	5432	0.08	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052021\
Data File : VU043839.D
Acq On : 19 May 2021 20:14
Operator : SY/MD
Sample : M2464-05
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
GB3Y8

Quant Time: May 20 03:57:29 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR051521WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu May 20 03:54:53 2021
Response via : Initial Calibration

