

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052021\
 Data File : VU043849.D
 Acq On : 20 May 2021 00:17
 Operator : SY/MD
 Sample : M2459-06 50X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0GK8

Quant Time: May 20 03:58:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR051521WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu May 20 03:54:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.260	114	202585	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.423	117	207667	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.819	152	94681	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	63042	4.22	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	84.40%
7) Chloroethane-d5	1.919	69	45840	4.39	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	87.80%
11) 1,1-Dichloroethene-d2	2.572	65	28382	4.58	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	91.60%
20) 2-Butanone-d5	4.645	46	168055	59.17	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	118.34%
24) Chloroform-d	5.073	84	124876	5.14	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.80%
26) 1,2-Dichloroethane-d4	5.713	65	63056	5.07	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.40%
32) Benzene-d6	5.735	84	257701	4.89	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.80%
36) 1,2-Dichloropropane-d6	6.700	67	83445	4.97	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.40%
41) Toluene-d8	7.906	98	213964	4.39	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.80%
43) trans-1,3-Dichloroprop...	8.189	79	26508	4.25	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.00%
46) 2-Hexanone-d5	8.645	63	127063	45.21	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	90.42%
56) 1,1,2,2-Tetrachloroeth...	10.764	84	66627	4.83	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.60%
66) 1,2-Dichlorobenzene-d4	12.201	152	84157	5.33	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	106.60%
Target Compounds						Qvalue
16) Methylene chloride	3.054	84	4305	0.22	ug/L	98
18) trans-1,2-Dichloroethene	3.363	96	1644	0.12	ug/L	95
22) cis-1,2-Dichloroethene	4.678	96	73762	5.00	ug/L	98
25) Chloroform	5.099	83	21883	0.87	ug/L	99
31) Carbon tetrachloride	5.533	117	50045	2.77	ug/L	96
34) Trichloroethene	6.549	95	463229	30.05	ug/L	97
47) Tetrachloroethene	8.562	164	108524	8.71	ug/L	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052021\
Data File : VU043849.D
Acq On : 20 May 2021 00:17
Operator : SY/MD
Sample : M2459-06 50X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0GK8

Quant Time: May 20 03:58:28 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR051521WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu May 20 03:54:53 2021
Response via : Initial Calibration









#16
Methylene chloride
Concen: 0.22 ug/L
RT: 3.054 min Scan# 533
Delta R.T. -0.000 min
Lab File: VU043849.D
Acq: 20 May 2021 00:17




Instrument :
MSVOA_U
ClientSampleId :
C0GK8


Tgt Ion:	84	Resp:	4305
Ion	Ratio	Lower	Upper
84	100		
86	61.5	44.7	82.9
49	116.6	83.1	154.3



#18
trans-1,2-Dichloroethene
Concen: 0.12 ug/L
RT: 3.363 min Scan# 629
Delta R.T. -0.000 min
Lab File: VU043849.D
Acq: 20 May 2021 00:17

Tgt Ion:	96	Resp:	1644
Ion	Ratio	Lower	Upper
96	100		
61	138.6	99.6	185.0
98	54.1	43.8	81.4









#22
cis-1,2-Dichloroethene
Concen: 5.00 ug/L
RT: 4.678 min Scan# 1038
Delta R.T. 0.000 min
Lab File: VU043849.D
Acq: 20 May 2021 00:17




Instrument :
MSVOA_U
ClientSampled :
C0GK8


Tgt Ion: 96 Resp: 73762
Ion Ratio Lower Upper
96 100
61 124.8 89.0 165.4
98 64.6 44.9 83.3



#25
Chloroform
Concen: 0.87 ug/L
RT: 5.099 min Scan# 1169
Delta R.T. -0.000 min
Lab File: VU043849.D
Acq: 20 May 2021 00:17

Tgt Ion: 83 Resp: 21883
Ion Ratio Lower Upper
83 100
85 63.7 45.0 83.6









#31
Carbon tetrachloride
Concen: 2.77 ug/L
RT: 5.533 min Scan# 1304
Delta R.T. -0.000 min
Lab File: VU043849.D
Acq: 20 May 2021 00:17




Instrument :
MSVOA_U
ClientSampleId :
C0GK8


Tgt Ion:117 Resp: 50045
Ion Ratio Lower Upper
117 100
119 100.7 77.4 116.0



#34
Trichloroethene
Concen: 30.05 ug/L
RT: 6.549 min Scan# 1620
Delta R.T. -0.003 min
Lab File: VU043849.D
Acq: 20 May 2021 00:17

Tgt Ion: 95 Resp: 463229
Ion Ratio Lower Upper
95 100
97 63.3 44.3 82.3
132 98.5 66.0 122.6
130 102.3 69.2 128.6





#47
Tetrachloroethene
Concen: 8.71 ug/L
RT: 8.562 min Scan# 2246
Delta R.T. -0.000 min
Lab File: VU043849.D
Acq: 20 May 2021 00:17

Instrument :
MSVOA_U
ClientSampleId :
C0GK8

Tgt Ion:164 Resp: 108524

Ion	Ratio	Lower	Upper
164	100		
129	89.1	64.1	119.1
131	87.3	61.2	113.8
166	124.9	87.8	163.2

