

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052022\
 Data File : VU048644.D
 Acq On : 20 May 2022 12:16
 Operator : SY/MD
 Sample : VU0520WBL01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK135

Quant Time: May 21 04:38:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM051622WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat May 21 04:37:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	219437	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	220885	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.815	152	124555	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	46633	34.696	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	69.400%
7) Chloroethane-d5	1.912	69	42487	42.194	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	84.380%
11) 1,1-Dichloroethene-d2	2.568	63	96313	31.686	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	63.380%
21) 2-Butanone-d5	4.629	46	128469	100.443	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	100.440%
24) Chloroform-d	5.067	84	162067	47.718	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.440%
26) 1,2-Dichloroethane-d4	5.703	65	121200	50.582	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.160%
32) Benzene-d6	5.729	84	278102	45.239	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.480%
36) 1,2-Dichloropropane-d6	6.693	67	82269	45.586	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	91.180%
41) Toluene-d8	7.899	98	264358	45.719	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.440%
43) trans-1,3-Dichloroprop...	8.182	79	50247	43.541	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.080%
47) 2-Hexanone-d5	8.636	63	89389	90.318	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	90.320%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	136066	45.564	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.120%
66) 1,2-Dichlorobenzene-d4	12.195	152	123847	48.639	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.280%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052022\
Data File : VU048644.D
Acq On : 20 May 2022 12:16
Operator : SY/MD
Sample : VU0520WBL01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK135

Quant Time: May 21 04:38:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM051622WMA.M
Quant Title : VOC Analysis
QLast Update : Sat May 21 04:37:22 2022
Response via : Initial Calibration

