

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052022\
 Data File : VU048647.D
 Acq On : 20 May 2022 16:47
 Operator : SY/MD
 Sample : N2926-21
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK002

Quant Time: May 21 04:38:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM051622WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat May 21 04:37:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	204188	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	205698	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	115281	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	42817	34.236	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	68.480%
7) Chloroethane-d5	1.912	69	37139	39.638	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	79.280%
11) 1,1-Dichloroethene-d2	2.568	63	89633	31.690	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	63.380%
21) 2-Butanone-d5	4.623	46	121572	102.149	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	102.150%
24) Chloroform-d	5.067	84	155056	49.063	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.120%
26) 1,2-Dichloroethane-d4	5.703	65	114630	51.413	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.820%
32) Benzene-d6	5.729	84	261536	45.685	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.380%
36) 1,2-Dichloropropane-d6	6.694	67	78602	46.770	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	93.540%
41) Toluene-d8	7.899	98	246041	45.692	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.380%
43) trans-1,3-Dichloroprop...	8.182	79	48980	45.577	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	91.160%
47) 2-Hexanone-d5	8.632	63	86909	94.295	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	94.300%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	135122	48.588	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	97.180%
66) 1,2-Dichlorobenzene-d4	12.195	152	119892	50.873	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.740%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052022\
Data File : VU048647.D
Acq On : 20 May 2022 16:47
Operator : SY/MD
Sample : N2926-21
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK002

Quant Time: May 21 04:38:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM051622WMA.M
Quant Title : VOC Analysis
QLast Update : Sat May 21 04:37:22 2022
Response via : Initial Calibration

