

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052024\
 Data File : VU058915.D
 Acq On : 20 May 2024 12:27
 Operator : MD/SY
 Sample : VSTD00217
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VSTD002017

Quant Time: May 21 02:39:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM052024.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue May 21 02:37:43 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.258	114	4196	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.418	117	6899	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.808	152	2593	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.597	65	6650	3.373	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.685	65	8559	2.376	ug/L	-0.02
7) 1,2-Dichloropropane-d6	6.687	67	7005	1.873	ug/L	0.00
8) Toluene-d8	7.887	98	18499	1.941	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.745	84	6603	2.402	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.597	62	8881	3.374	ug/L	93
6) Trichloroethene	6.544	95	6775	1.627	ug/L	98
9) 1,2-Dibromoethane	8.924	107	4940	2.391	ug/L #	95
12) 1,2,3-Trichloropropane	10.797	75	5190	2.134	ug/L	99
13) 1,2-Dibromo-3-chloropr...	12.973	75	1235	2.511	ug/L #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU052024\
Data File : VU058915.D
Acq On : 20 May 2024 12:27
Operator : MD/SY
Sample : VSTD00217
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTD002017

Quant Time: May 21 02:39:05 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM052024.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue May 21 02:37:43 2024
Response via : Initial Calibration

