

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU052119\
 Data File : VU032194.D
 Acq On : 21 May 2019 12:59
 Operator : JC/SP
 Sample : K2788-14ME
 Misc : 5.30g/5mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 A51A4ME

Manual Integrations
 APPROVED

MMDadoda
 5/22/2019 10:04:32 AM

Quant Time: May 22 02:22:57 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM052119WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed May 22 02:08:11 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	862698	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	852258	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	400748	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	225177	40.02	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	80.04%
7) Chloroethane-d5	1.64	69	109191	22.27	ug/L	-0.03
Spiked Amount	50.000	Range	70 - 130	Recovery	=	44.54%#
11) 1,1-Dichloroethene-d2	2.21	63	335277m	31.29	ug/L	-0.05
Spiked Amount	50.000	Range	60 - 125	Recovery	=	62.58%
21) 2-Butanone-d5	4.21	46	330823	95.28	ug/L	0.04
Spiked Amount	100.000	Range	40 - 130	Recovery	=	95.28%
24) Chloroform-d	4.64	84	452208	40.78	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	81.56%
26) 1,2-Dichloroethane-d4	5.30	65	284260	43.43	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.86%
32) Benzene-d6	5.32	84	961279	43.39	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.78%
36) 1,2-Dichloropropane-d6	6.32	67	286043	44.72	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	89.44%
41) Toluene-d8	7.56	98	901607	42.94	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	85.88%
43) trans-1,3-Dichloropropene-	7.85	79	133337	42.24	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	84.48%
47) 2-Hexanone-d5	8.33	63	271445	93.54	ug/L	0.02
Spiked Amount	100.000	Range	45 - 130	Recovery	=	93.54%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	462228	44.67	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.34%
64) 1,2-Dichlorobenzene-d4	11.86	152	364301	43.46	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	86.92%

Target Compounds

					Ovalue
20) cis-1,2-Dichloroethene	4.21	96	741336	110.834	ug/L 99
30) 1,1,1-Trichloroethane	4.90	97	8576	0.943	ug/L # 79
34) Trichloroethene	6.17	95	1665186	261.697	ug/L 98
46) Tetrachloroethene	8.22	164	25014	4.761	ug/L 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU052119\
Data File : VU032194.D
Acq On : 21 May 2019 12:59
Operator : JC/SP
Sample : K2788-14ME
Misc : 5.30g/5mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
A51A4ME

Manual Integrations
APPROVED

MMDadoda
5/22/2019 10:04:32 AM

Quant Time: May 22 02:22:57 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM052119WMA.M
Quant Title : VOC Analysis
QLast Update : Wed May 22 02:08:11 2019
Response via : Initial Calibration

