

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU052219\
 Data File : VU032240.D
 Acq On : 22 May 2019 22:31
 Operator : JC/SP
 Sample : K2888-10
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0BP2

Quant Time: May 23 09:42:41 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR052219WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu May 23 01:46:44 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	382205	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	398917	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	155409	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	113193	5.27	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	105.40%
7) Chloroethane-d5	1.68	69	97889	4.89	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	97.80%
11) 1,1-Dichloroethene-d2	2.27	63	169861	4.29	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	85.80%
20) 2-Butanone-d5	4.20	46	251061	41.47	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	82.94%
24) Chloroform-d	4.64	84	224006	4.09	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.80%
26) 1,2-Dichloroethane-d4	5.31	65	111146	3.83	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	76.60%
32) Benzene-d6	5.33	84	458319	4.05	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.00%
36) 1,2-Dichloropropane-d6	6.33	67	139330	3.79	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	75.80%
41) Toluene-d8	7.56	98	381235	3.47	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	69.40%#
43) trans-1,3-Dichloropropene-	7.85	79	42907	3.04	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	60.80%
46) 2-Hexanone-d5	8.31	63	240364	45.19	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	90.38%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	104749	2.87	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	57.40%#
64) 1,2-Dichlorobenzene-d4	11.86	152	140108	3.27	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	65.40%#

Target Compounds					Ovalue
13) Acetone	2.34	43	9957	5.308 ug/L	98
47) Tetrachloroethene	8.22	164	4600941	384.840 ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU052219\
Data File : VU032240.D
Acq On : 22 May 2019 22:31
Operator : JC/SP
Sample : K2888-10
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0BP2

Quant Time: May 23 09:42:41 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR052219WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu May 23 01:46:44 2019
Response via : Initial Calibration

